

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: EarthEfficient LLC

SDG No.: P5217

Contract: EART12

Lab Code: CHEM

Case No.: P5217

SAS No.: P5217

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165528BS Lead	mg/Kg	42.6	41.7		98	80 - 120	P