

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB165518BS

Contract: EART12

Lab Code: CHEM Case No.: P5217 SAS No.: P5217 SDG NO.: P5217

Lab Sample ID: PB165518BS Date(s) Analyzed: 12/10/2024 12/10/2024

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO108437.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.806	4.756	4.856	156	154	1.31
COLUMN 1	2	4.827	4.777	4.877	156		
	3	4.882	4.832	4.932	155		
	4	5.004	4.954	5.054	155		
	5	5.261	5.211	5.311	150		
	1	4.794	4.744	4.844	159		
COLUMN 2	2	4.814	4.764	4.864	159	152	
	3	4.989	4.939	5.039	157		
	4	5.031	4.981	5.081	154		
	5	5.245	5.195	5.295	132		
Aroclor-1260	1	6.306	6.256	6.356	160	146	9.77
COLUMN 1	2	6.494	6.444	6.544	160		
	3	6.863	6.813	6.913	139		
	4	7.126	7.076	7.176	136		
	5	7.366	7.316	7.416	134		
	1	6.28	6.23	6.33	167		
COLUMN 2	2	6.467	6.417	6.517	166	161	
	3	6.621	6.571	6.671	167		
	4	7.094	7.044	7.144	150		
	5	7.333	7.283	7.383	152		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

BP-G-6MS

Contract: EART12

Lab Code: CHEM

Case No.: P5217

SAS No.: P5217

SDG NO.: P5217

Lab Sample ID: P5216-01MS

Date(s) Analyzed: 12/10/2024

12/10/2024

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO108440.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.807	4.757	4.857	206	204	3.37
COLUMN 1	2	4.827	4.777	4.877	208		
	3	4.883	4.833	4.933	206		
	4	5.004	4.954	5.054	204		
	5	5.262	5.212	5.312	197		
	1	4.794	4.744	4.844	218		
COLUMN 2	2	4.814	4.764	4.864	217	211	
	3	4.99	4.94	5.04	211		
	4	5.031	4.981	5.081	204		
	5	5.245	5.195	5.295	205		
Aroclor-1260	1	6.306	6.256	6.356	203	183	12.79
COLUMN 1	2	6.494	6.444	6.544	199		
	3	6.864	6.814	6.914	173		
	4	7.125	7.075	7.175	174		
	5	7.366	7.316	7.416	168		
	1	6.281	6.231	6.331	219		
COLUMN 2	2	6.467	6.417	6.517	215	208	
	3	6.621	6.571	6.671	216		
	4	7.094	7.044	7.144	194		
	5	7.333	7.283	7.383	197		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

BP-G-6MSD

Contract: EART12

Lab Code: CHEM Case No.: P5217 SAS No.: P5217 SDG NO.: P5217

Lab Sample ID: P5216-01MSD Date(s) Analyzed: 12/10/2024 12/10/2024

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO108441.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.807	4.757	4.857	206	203	3.86
COLUMN 1	2	4.827	4.777	4.877	206		
	3	4.882	4.832	4.932	204		
	4	5.004	4.954	5.054	203		
	5	5.262	5.212	5.312	196		
	1	4.794	4.744	4.844	218		
COLUMN 2	2	4.814	4.764	4.864	219	211	
	3	4.99	4.94	5.04	212		
	4	5.032	4.982	5.082	204		
	5	5.245	5.195	5.295	204		
Aroclor-1260	1	6.306	6.256	6.356	203	185	11.7
COLUMN 1	2	6.494	6.444	6.544	201		
	3	6.865	6.815	6.915	173		
	4	7.125	7.075	7.175	177		
	5	7.366	7.316	7.416	173		
	1	6.281	6.231	6.331	219		
COLUMN 2	2	6.468	6.418	6.518	215	208	
	3	6.621	6.571	6.671	217		
	4	7.094	7.044	7.144	194		
	5	7.333	7.283	7.383	197		