

CALIBRATION VERIFICATION SUMMARY

Contract: EART12

Lab Code: CHEM Case No.: P5217 SAS No.: P5217 SDG NO.: P5217

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/10/2024

Lab Sample No.: AR1660CCC500 Data File : PO108430.D Time Analyzed: 09:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.806	4.709	4.909	528.530	500.000	5.7
Aroclor-1016-2	4.826	4.729	4.929	535.930	500.000	7.2
Aroclor-1016-3	4.882	4.785	4.985	527.970	500.000	5.6
Aroclor-1016-4	5.003	4.907	5.107	539.570	500.000	7.9
Aroclor-1016-5	5.262	5.165	5.365	540.020	500.000	8.0
Aroclor-1260-1	6.306	6.210	6.410	518.530	500.000	3.7
Aroclor-1260-2	6.494	6.398	6.598	512.820	500.000	2.6
Aroclor-1260-3	6.864	6.769	6.969	480.430	500.000	-3.9
Aroclor-1260-4	7.125	7.029	7.229	463.890	500.000	-7.2
Aroclor-1260-5	7.366	7.270	7.470	497.800	500.000	-0.4
Decachlorobiphenyl	8.784	8.691	8.891	45.050	50.000	-9.9
Tetrachloro-m-xylene	3.708	3.610	3.810	53.830	50.000	7.7

CALIBRATION VERIFICATION SUMMARY

Contract: EART12

Lab Code: CHEM Case No.: P5217 SAS No.: P5217 SDG NO.: P5217

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/10/2024

Lab Sample No.: AR1660CCC500 Data File : PO108430.D Time Analyzed: 09:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.794	4.698	4.898	549.880	500.000	10.0
Aroclor-1016-2	4.814	4.718	4.918	546.860	500.000	9.4
Aroclor-1016-3	4.990	4.894	5.094	534.700	500.000	6.9
Aroclor-1016-4	5.031	4.935	5.135	491.320	500.000	-1.7
Aroclor-1016-5	5.245	5.150	5.350	501.850	500.000	0.4
Aroclor-1260-1	6.281	6.186	6.386	545.810	500.000	9.2
Aroclor-1260-2	6.468	6.373	6.573	554.750	500.000	11.0
Aroclor-1260-3	6.622	6.527	6.727	544.860	500.000	9.0
Aroclor-1260-4	7.094	7.000	7.200	541.050	500.000	8.2
Aroclor-1260-5	7.334	7.239	7.439	543.290	500.000	8.7
Decachlorobiphenyl	8.732	8.641	8.841	50.360	50.000	0.7
Tetrachloro-m-xylene	3.706	3.608	3.808	54.120	50.000	8.2