

CALIBRATION VERIFICATION SUMMARY

Contract: EART12

Lab Code: CHEM Case No.: P5217 SAS No.: P5217 SDG NO.: P5217

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/10/2024

Lab Sample No.: AR1660CCC500 Data File : PO108445.D Time Analyzed: 15:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.807	4.709	4.909	516.500	500.000	3.3
Aroclor-1016-2	4.826	4.729	4.929	526.670	500.000	5.3
Aroclor-1016-3	4.882	4.785	4.985	523.180	500.000	4.6
Aroclor-1016-4	5.004	4.907	5.107	521.190	500.000	4.2
Aroclor-1016-5	5.262	5.165	5.365	515.310	500.000	3.1
Aroclor-1260-1	6.306	6.210	6.410	505.860	500.000	1.2
Aroclor-1260-2	6.494	6.398	6.598	507.010	500.000	1.4
Aroclor-1260-3	6.863	6.769	6.969	509.420	500.000	1.9
Aroclor-1260-4	7.124	7.029	7.229	505.260	500.000	1.1
Aroclor-1260-5	7.365	7.270	7.470	505.220	500.000	1.0
Decachlorobiphenyl	8.781	8.691	8.891	47.940	50.000	-4.1
Tetrachloro-m-xylene	3.709	3.610	3.810	52.390	50.000	4.8

CALIBRATION VERIFICATION SUMMARY

Contract: EART12

Lab Code: CHEM Case No.: P5217 SAS No.: P5217 SDG NO.: P5217

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/10/2024

Lab Sample No.: AR1660CCC500 Data File : PO108445.D Time Analyzed: 15:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.794	4.698	4.898	552.000	500.000	10.4
Aroclor-1016-2	4.814	4.718	4.918	552.380	500.000	10.5
Aroclor-1016-3	4.990	4.894	5.094	540.910	500.000	8.2
Aroclor-1016-4	5.031	4.935	5.135	540.230	500.000	8.0
Aroclor-1016-5	5.245	5.150	5.350	493.430	500.000	-1.3
Aroclor-1260-1	6.280	6.186	6.386	549.290	500.000	9.9
Aroclor-1260-2	6.467	6.373	6.573	551.630	500.000	10.3
Aroclor-1260-3	6.621	6.527	6.727	547.660	500.000	9.5
Aroclor-1260-4	7.094	7.000	7.200	564.580	500.000	12.9
Aroclor-1260-5	7.333	7.239	7.439	570.670	500.000	14.1
Decachlorobiphenyl	8.732	8.641	8.841	52.580	50.000	5.2
Tetrachloro-m-xylene	3.706	3.608	3.808	54.660	50.000	9.3