



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

ENTACT

Project Name: North Point

Project # N/A

Chemtech Project # P5224

Test Name: TCLP Metals Group2, TCLP Mercury

A. Number of Samples and Date of Receipt:

2 Water samples were received on 12/09/2024.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: PCB, pH, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP Metals Group2 and VOC-TCLVOA-10. This data package contains results for TCLP Metals Group2, TCLP Mercury.

C. Analytical Techniques:

The analysis of TCLP Metals Group2 was based on method 6010D, digestion based on method 3010 (waters). The analysis and digestion of TCLP Mercury was based on method 7470A and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike (NP-WS-001MS) analysis met criteria for all samples except for Mercury due to matrix interference.

The Matrix Spike Duplicate (NP-WS-001MSD) analysis met criteria for all samples except for Mercury due to matrix interference.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

FAX and Hardcopy Not match for Sample P5224-02 as During FAX time, its Qcs Samples not analyzed in sequences , so sample re-analyze with its Qcs in another sequence , so we reported correct data in Hardcopy.



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____