



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** P5224
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P5224 **SAS No.:** P5224
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.24	0.2	118	40 - 160	CV	12/13/2024	10:35	LB133933
CRI01	Aluminum	94.7	100	95	40 - 160	P	12/16/2024	16:14	LB133968
	Arsenic	18.9	20.0	95	40 - 160	P	12/16/2024	16:14	LB133968
	Barium	97.7	100	98	40 - 160	P	12/16/2024	16:14	LB133968
	Cadmium	5.61	6.0	94	40 - 160	P	12/16/2024	16:14	LB133968
	Chromium	9.95	10.0	100	40 - 160	P	12/16/2024	16:14	LB133968
	Cobalt	28.5	30.0	95	40 - 160	P	12/16/2024	16:14	LB133968
	Copper	21.3	20.0	107	40 - 160	P	12/16/2024	16:14	LB133968
	Iron	91.4	100	91	40 - 160	P	12/16/2024	16:14	LB133968
	Lead	12.0	12.0	100	40 - 160	P	12/16/2024	16:14	LB133968
	Nickel	38.0	40.0	95	40 - 160	P	12/16/2024	16:14	LB133968
	Selenium	19.0	20.0	95	40 - 160	P	12/16/2024	16:14	LB133968
	Silver	9.77	10.0	98	40 - 160	P	12/16/2024	16:14	LB133968
	Zinc	43.5	40.0	109	40 - 160	P	12/16/2024	16:14	LB133968