

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** P5224
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P5224 **SAS No.:** P5224

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165597BS							
Aluminum	ug/L	10000	9930		99	80 - 120	P
Arsenic	ug/L	4000	3750		94	80 - 120	P
Barium	ug/L	1000	976		98	80 - 120	P
Cadmium	ug/L	1000	992		99	80 - 120	P
Chromium	ug/L	2000	2060		103	80 - 120	P
Cobalt	ug/L	1000	997		100	80 - 120	P
Copper	ug/L	1500	1530		102	80 - 120	P
Iron	ug/L	15000	14600		97	80 - 120	P
Lead	ug/L	5000	4940		99	80 - 120	P
Nickel	ug/L	2500	2510		100	80 - 120	P
Selenium	ug/L	10000	9480		95	80 - 120	P
Silver	ug/L	380	372		98	80 - 120	P
Zinc	ug/L	1000	1020		102	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: P5224

Contract: ENTA05

Lab Code: CHEM

Case No.: P5224

SAS No.: P5224

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165599BS Mercury	ug/L	4.0	3.62		90	80 - 120	CV