

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P5224 SAS No.: P5224 SDG NO.: P5224

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/10/2024

Lab Sample No.: AR1660CCC500 Data File : PO108460.D Time Analyzed: 21:10

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|----------------------|-------|--------------------|-------------------|-------|
| Aroclor-1016-1       | 4.806 | 4.709                | 4.909 | 520.980            | 500.000           | 4.2   |
| Aroclor-1016-2       | 4.826 | 4.729                | 4.929 | 517.570            | 500.000           | 3.5   |
| Aroclor-1016-3       | 4.882 | 4.785                | 4.985 | 519.270            | 500.000           | 3.9   |
| Aroclor-1016-4       | 5.004 | 4.907                | 5.107 | 517.350            | 500.000           | 3.5   |
| Aroclor-1016-5       | 5.262 | 5.165                | 5.365 | 500.070            | 500.000           | 0.0   |
| Aroclor-1260-1       | 6.305 | 6.210                | 6.410 | 472.020            | 500.000           | -5.6  |
| Aroclor-1260-2       | 6.494 | 6.398                | 6.598 | 452.680            | 500.000           | -9.5  |
| Aroclor-1260-3       | 6.863 | 6.769                | 6.969 | 449.120            | 500.000           | -10.2 |
| Aroclor-1260-4       | 7.124 | 7.029                | 7.229 | 460.950            | 500.000           | -7.8  |
| Aroclor-1260-5       | 7.365 | 7.270                | 7.470 | 453.630            | 500.000           | -9.3  |
| Decachlorobiphenyl   | 8.782 | 8.691                | 8.891 | 41.980             | 50.000            | -16.0 |
| Tetrachloro-m-xylene | 3.709 | 3.610                | 3.810 | 52.250             | 50.000            | 4.5   |

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P5224 SAS No.: P5224 SDG NO.: P5224

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/10/2024

Lab Sample No.: AR1660CCC500 Data File : PO108460.D Time Analyzed: 21:10

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.794 | 4.698     | 4.898 | 569.110            | 500.000           | 13.8 |
| Aroclor-1016-2       | 4.814 | 4.718     | 4.918 | 562.750            | 500.000           | 12.6 |
| Aroclor-1016-3       | 4.990 | 4.894     | 5.094 | 555.150            | 500.000           | 11.0 |
| Aroclor-1016-4       | 5.032 | 4.935     | 5.135 | 543.960            | 500.000           | 8.8  |
| Aroclor-1016-5       | 5.246 | 5.150     | 5.350 | 503.000            | 500.000           | 0.6  |
| Aroclor-1260-1       | 6.281 | 6.186     | 6.386 | 539.260            | 500.000           | 7.9  |
| Aroclor-1260-2       | 6.467 | 6.373     | 6.573 | 527.630            | 500.000           | 5.5  |
| Aroclor-1260-3       | 6.621 | 6.527     | 6.727 | 516.760            | 500.000           | 3.4  |
| Aroclor-1260-4       | 7.094 | 7.000     | 7.200 | 529.440            | 500.000           | 5.9  |
| Aroclor-1260-5       | 7.333 | 7.239     | 7.439 | 526.350            | 500.000           | 5.3  |
| Decachlorobiphenyl   | 8.732 | 8.641     | 8.841 | 49.970             | 50.000            | -0.1 |
| Tetrachloro-m-xylene | 3.707 | 3.608     | 3.808 | 56.150             | 50.000            | 12.3 |