

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P5224 SAS No.: P5224 SDG NO.: P5224

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL03 Date Analyzed: 12/11/2024

Lab Sample No.: AR1660CCC500 Data File : PO108475.D Time Analyzed: 02:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.804	4.709	4.909	530.590	500.000	6.1
Aroclor-1016-2	4.824	4.729	4.929	533.640	500.000	6.7
Aroclor-1016-3	4.880	4.785	4.985	530.880	500.000	6.2
Aroclor-1016-4	5.001	4.907	5.107	537.220	500.000	7.4
Aroclor-1016-5	5.259	5.165	5.365	530.630	500.000	6.1
Aroclor-1260-1	6.303	6.210	6.410	510.350	500.000	2.1
Aroclor-1260-2	6.492	6.398	6.598	507.720	500.000	1.5
Aroclor-1260-3	6.861	6.769	6.969	504.670	500.000	0.9
Aroclor-1260-4	7.122	7.029	7.229	492.130	500.000	-1.6
Aroclor-1260-5	7.364	7.270	7.470	497.830	500.000	-0.4
Decachlorobiphenyl	8.781	8.691	8.891	45.960	50.000	-8.1
Tetrachloro-m-xylene	3.706	3.610	3.810	53.300	50.000	6.6

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P5224 SAS No.: P5224 SDG NO.: P5224

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL03 Date Analyzed: 12/11/2024

Lab Sample No.: AR1660CCC500 Data File : PO108475.D Time Analyzed: 02:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.794	4.698	4.898	578.270	500.000	15.7
Aroclor-1016-2	4.814	4.718	4.918	579.530	500.000	15.9
Aroclor-1016-3	4.990	4.894	5.094	565.940	500.000	13.2
Aroclor-1016-4	5.031	4.935	5.135	542.840	500.000	8.6
Aroclor-1016-5	5.245	5.150	5.350	510.680	500.000	2.1
Aroclor-1260-1	6.280	6.186	6.386	569.060	500.000	13.8
Aroclor-1260-2	6.467	6.373	6.573	570.030	500.000	14.0
Aroclor-1260-3	6.621	6.527	6.727	565.470	500.000	13.1
Aroclor-1260-4	7.093	7.000	7.200	578.900	500.000	15.8
Aroclor-1260-5	7.332	7.239	7.439	584.100	500.000	16.8
Decachlorobiphenyl	8.731	8.641	8.841	54.570	50.000	9.1
Tetrachloro-m-xylene	3.706	3.608	3.808	57.140	50.000	14.3