

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5225 SAS No.: P5225 SDG NO.: P5225

Continuing Calib Date: 12/11/2024 Initial Calibration Date(s): 12/06/2024 12/06/2024

Continuing Calib Time: 17:43 Initial Calibration Time(s): 14:19 22:34

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.81	4.81	4.71	4.91	0.00
Aroclor-1248-2 (2)	5.05	5.05	4.95	5.15	0.00
Aroclor-1248-3 (3)	5.26	5.26	5.16	5.36	0.00
Aroclor-1248-4 (4)	5.62	5.62	5.52	5.72	0.00
Aroclor-1248-5 (5)	5.66	5.66	5.56	5.76	0.00
Tetrachloro-m-xylene	3.71	3.71	3.61	3.81	0.00
Decachlorobiphenyl	8.78	8.79	8.69	8.89	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM **Case No.:** P5225 **SAS No.:** P5225 **SDG NO.:** P5225

Continuing Calib Date: 12/11/2024 **Initial Calibration Date(s):** 12/06/2024 12/06/2024

Continuing Calib Time: 17:43 **Initial Calibration Time(s):** 14:19 22:34

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.79	4.80	4.70	4.90	0.01
Aroclor-1248-2 (2)	5.03	5.03	4.93	5.13	0.00
Aroclor-1248-3 (3)	5.08	5.08	4.98	5.18	0.01
Aroclor-1248-4 (4)	5.25	5.25	5.15	5.35	0.00
Aroclor-1248-5 (5)	5.64	5.64	5.54	5.74	0.00
Tetrachloro-m-xylene	3.71	3.71	3.61	3.81	0.00
Decachlorobiphenyl	8.73	8.74	8.64	8.84	0.01