

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5225 SAS No.: P5225 SDG NO.: P5225

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/11/2024

Lab Sample No.: AR1660CCC500 Data File : PO108504.D Time Analyzed: 17:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.807	4.709	4.909	529.860	500.000	6.0
Aroclor-1016-2	4.827	4.729	4.929	535.680	500.000	7.1
Aroclor-1016-3	4.883	4.785	4.985	534.500	500.000	6.9
Aroclor-1016-4	5.004	4.907	5.107	538.530	500.000	7.7
Aroclor-1016-5	5.262	5.165	5.365	526.090	500.000	5.2
Aroclor-1260-1	6.307	6.210	6.410	483.830	500.000	-3.2
Aroclor-1260-2	6.495	6.398	6.598	480.620	500.000	-3.9
Aroclor-1260-3	6.865	6.769	6.969	460.250	500.000	-8.0
Aroclor-1260-4	7.126	7.029	7.229	471.060	500.000	-5.8
Aroclor-1260-5	7.367	7.270	7.470	459.570	500.000	-8.1
Decachlorobiphenyl	8.786	8.691	8.891	42.320	50.000	-15.4
Tetrachloro-m-xylene	3.709	3.610	3.810	53.240	50.000	6.5

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5225 SAS No.: P5225 SDG NO.: P5225

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/11/2024

Lab Sample No.: AR1660CCC500 Data File : PO108504.D Time Analyzed: 17:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.796	4.698	4.898	586.740	500.000	17.3
Aroclor-1016-2	4.815	4.718	4.918	586.700	500.000	17.3
Aroclor-1016-3	4.991	4.894	5.094	578.750	500.000	15.8
Aroclor-1016-4	5.032	4.935	5.135	556.140	500.000	11.2
Aroclor-1016-5	5.246	5.150	5.350	547.000	500.000	9.4
Aroclor-1260-1	6.282	6.186	6.386	556.840	500.000	11.4
Aroclor-1260-2	6.469	6.373	6.573	547.980	500.000	9.6
Aroclor-1260-3	6.623	6.527	6.727	532.800	500.000	6.6
Aroclor-1260-4	7.096	7.000	7.200	546.060	500.000	9.2
Aroclor-1260-5	7.336	7.239	7.439	541.410	500.000	8.3
Decachlorobiphenyl	8.735	8.641	8.841	51.380	50.000	2.8
Tetrachloro-m-xylene	3.707	3.608	3.808	57.720	50.000	15.4