

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5225 SAS No.: P5225 SDG NO.: P5225

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/11/2024

Lab Sample No.: AR1660CCC500 Data File : PO108515.D Time Analyzed: 21:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.806	4.709	4.909	523.150	500.000	4.6
Aroclor-1016-2	4.826	4.729	4.929	525.350	500.000	5.1
Aroclor-1016-3	4.882	4.785	4.985	519.630	500.000	3.9
Aroclor-1016-4	5.003	4.907	5.107	523.000	500.000	4.6
Aroclor-1016-5	5.261	5.165	5.365	515.760	500.000	3.2
Aroclor-1260-1	6.305	6.210	6.410	492.900	500.000	-1.4
Aroclor-1260-2	6.494	6.398	6.598	488.300	500.000	-2.3
Aroclor-1260-3	6.863	6.769	6.969	473.230	500.000	-5.4
Aroclor-1260-4	7.124	7.029	7.229	465.030	500.000	-7.0
Aroclor-1260-5	7.365	7.270	7.470	475.870	500.000	-4.8
Decachlorobiphenyl	8.782	8.691	8.891	45.250	50.000	-9.5
Tetrachloro-m-xylene	3.708	3.610	3.810	52.960	50.000	5.9

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5225 SAS No.: P5225 SDG NO.: P5225

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/11/2024

Lab Sample No.: AR1660CCC500 Data File : PO108515.D Time Analyzed: 21:24

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.794	4.698	4.898	578.250	500.000	15.7
Aroclor-1016-2	4.814	4.718	4.918	577.870	500.000	15.6
Aroclor-1016-3	4.990	4.894	5.094	566.950	500.000	13.4
Aroclor-1016-4	5.031	4.935	5.135	546.780	500.000	9.4
Aroclor-1016-5	5.245	5.150	5.350	501.650	500.000	0.3
Aroclor-1260-1	6.280	6.186	6.386	551.980	500.000	10.4
Aroclor-1260-2	6.467	6.373	6.573	555.900	500.000	11.2
Aroclor-1260-3	6.622	6.527	6.727	557.320	500.000	11.5
Aroclor-1260-4	7.094	7.000	7.200	565.150	500.000	13.0
Aroclor-1260-5	7.333	7.239	7.439	568.990	500.000	13.8
Decachlorobiphenyl	8.732	8.641	8.841	53.860	50.000	7.7
Tetrachloro-m-xylene	3.706	3.608	3.808	56.980	50.000	14.0