

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5225 SAS No.: P5225 SDG NO.: P5225

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL03 Date Analyzed: 12/12/2024

Lab Sample No.: AR1660CCC500 Data File : PO108521.D Time Analyzed: 09:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.805	4.709	4.909	535.590	500.000	7.1
Aroclor-1016-2	4.825	4.729	4.929	540.290	500.000	8.1
Aroclor-1016-3	4.880	4.785	4.985	533.090	500.000	6.6
Aroclor-1016-4	5.001	4.907	5.107	540.030	500.000	8.0
Aroclor-1016-5	5.260	5.165	5.365	536.090	500.000	7.2
Aroclor-1260-1	6.303	6.210	6.410	514.540	500.000	2.9
Aroclor-1260-2	6.492	6.398	6.598	516.910	500.000	3.4
Aroclor-1260-3	6.862	6.769	6.969	520.350	500.000	4.1
Aroclor-1260-4	7.122	7.029	7.229	508.300	500.000	1.7
Aroclor-1260-5	7.363	7.270	7.470	506.430	500.000	1.3
Decachlorobiphenyl	8.780	8.691	8.891	49.280	50.000	-1.4
Tetrachloro-m-xylene	3.707	3.610	3.810	54.340	50.000	8.7

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5225 SAS No.: P5225 SDG NO.: P5225

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL03 Date Analyzed: 12/12/2024

Lab Sample No.: AR1660CCC500 Data File : PO108521.D Time Analyzed: 09:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.791	4.698	4.898	559.120	500.000	11.8
Aroclor-1016-2	4.811	4.718	4.918	560.570	500.000	12.1
Aroclor-1016-3	4.987	4.894	5.094	545.720	500.000	9.1
Aroclor-1016-4	5.028	4.935	5.135	526.760	500.000	5.4
Aroclor-1016-5	5.242	5.150	5.350	417.410	500.000	-16.5
Aroclor-1260-1	6.277	6.186	6.386	554.250	500.000	10.9
Aroclor-1260-2	6.465	6.373	6.573	556.540	500.000	11.3
Aroclor-1260-3	6.619	6.527	6.727	551.620	500.000	10.3
Aroclor-1260-4	7.091	7.000	7.200	570.930	500.000	14.2
Aroclor-1260-5	7.331	7.239	7.439	584.070	500.000	16.8
Decachlorobiphenyl	8.728	8.641	8.841	55.120	50.000	10.2
Tetrachloro-m-xylene	3.704	3.608	3.808	53.440	50.000	6.9