

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5225 SAS No.: P5225 SDG NO.: P5225

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL04 Date Analyzed: 12/12/2024

Lab Sample No.: AR1660CCC500 Data File : PO108530.D Time Analyzed: 16:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.805	4.709	4.909	481.860	500.000	-3.6
Aroclor-1016-2	4.824	4.729	4.929	480.950	500.000	-3.8
Aroclor-1016-3	4.880	4.785	4.985	480.260	500.000	-3.9
Aroclor-1016-4	5.002	4.907	5.107	509.320	500.000	1.9
Aroclor-1016-5	5.260	5.165	5.365	538.000	500.000	7.6
Aroclor-1260-1	6.303	6.210	6.410	423.100	500.000	-15.4
Aroclor-1260-2	6.491	6.398	6.598	429.780	500.000	-14.0
Aroclor-1260-3	6.861	6.769	6.969	419.250	500.000	-16.2
Aroclor-1260-4	7.122	7.029	7.229	412.370	500.000	-17.5
Aroclor-1260-5	7.362	7.270	7.470	414.110	500.000	-17.2
Decachlorobiphenyl	8.778	8.691	8.891	43.030	50.000	-13.9
Tetrachloro-m-xylene	3.707	3.610	3.810	52.380	50.000	4.8

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5225 SAS No.: P5225 SDG NO.: P5225

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL04 Date Analyzed: 12/12/2024

Lab Sample No.: AR1660CCC500 Data File : PO108530.D Time Analyzed: 16:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.792	4.698	4.898	535.660	500.000	7.1
Aroclor-1016-2	4.812	4.718	4.918	537.740	500.000	7.5
Aroclor-1016-3	4.988	4.894	5.094	523.010	500.000	4.6
Aroclor-1016-4	5.028	4.935	5.135	458.290	500.000	-8.3
Aroclor-1016-5	5.243	5.150	5.350	585.890	500.000	17.2
Aroclor-1260-1	6.278	6.186	6.386	497.420	500.000	-0.5
Aroclor-1260-2	6.465	6.373	6.573	514.530	500.000	2.9
Aroclor-1260-3	6.619	6.527	6.727	509.930	500.000	2.0
Aroclor-1260-4	7.091	7.000	7.200	509.590	500.000	1.9
Aroclor-1260-5	7.330	7.239	7.439	511.270	500.000	2.3
Decachlorobiphenyl	8.728	8.641	8.841	50.860	50.000	1.7
Tetrachloro-m-xylene	3.704	3.608	3.808	53.910	50.000	7.8