

Surrogate Summary

SDG No.: **P5225**

Client: **ATC Group Services LLC**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO108361.D	PIBLK-PO108361.D	Tetrachloro-m-xylene	1	20	23.0	115		10	157
		Decachlorobiphenyl	1	20	23.6	118		10	173
		Tetrachloro-m-xylene	2	20	21.6	108		10	157
		Decachlorobiphenyl	2	20	23.6	118		10	173
I.BLK-PO108508.D	PIBLK-PO108508.D	Tetrachloro-m-xylene	1	20	26.1	131		60	140
		Decachlorobiphenyl	1	20	23.4	117		60	140
		Tetrachloro-m-xylene	2	20	26.8	134		60	140
		Decachlorobiphenyl	2	20	27.9	140		60	140
PB165544BL	PB165544BL	Tetrachloro-m-xylene	1	20	21.8	109		32	144
		Decachlorobiphenyl	1	20	20.5	103		32	175
		Tetrachloro-m-xylene	2	20	22.3	112		32	144
		Decachlorobiphenyl	2	20	24.3	121		32	175
PB165544BS	PB165544BS	Tetrachloro-m-xylene	1	20	20.3	102		32	144
		Decachlorobiphenyl	1	20	20.8	104		32	175
		Tetrachloro-m-xylene	2	20	20.5	103		32	144
		Decachlorobiphenyl	2	20	24.4	122		32	175
P5225-01	1A-1B-1C-MB	Tetrachloro-m-xylene	1	20	20.4	102		32	144
		Decachlorobiphenyl	1	20	28.0	140		32	175
		Tetrachloro-m-xylene	2	20	21.2	106		32	144
		Decachlorobiphenyl	2	20	32.3	161		32	175
P5225-02	2A-2B-2C-MB	Tetrachloro-m-xylene	1	20	14.6	73		32	144
		Decachlorobiphenyl	1	20	31.7	159		32	175
		Tetrachloro-m-xylene	2	20	15.7	79		32	144
		Decachlorobiphenyl	2	20	36.1	181	*	32	175
P5225-03	3A-3B-3C-AB	Tetrachloro-m-xylene	1	20	20.4	102		32	144
		Decachlorobiphenyl	1	20	22.3	112		32	175
		Tetrachloro-m-xylene	2	20	21.0	105		32	144
		Decachlorobiphenyl	2	20	25.6	128		32	175
P5225-04	4A-4B-4C-BASIC	Tetrachloro-m-xylene	1	20	15.3	77		32	144
		Decachlorobiphenyl	1	20	69.1	346	*	32	175
		Tetrachloro-m-xylene	2	20	17.2	86		32	144
		Decachlorobiphenyl	2	20	78.7	393	*	32	175
I.BLK-PO108519.D	PIBLK-PO108519.D	Tetrachloro-m-xylene	1	20	24.4	122		60	140
		Decachlorobiphenyl	1	20	24.3	121		60	140
		Tetrachloro-m-xylene	2	20	25.1	126		60	140
		Decachlorobiphenyl	2	20	28.1	140		60	140
I.BLK-PO108522.D	PIBLK-PO108522.D	Tetrachloro-m-xylene	1	20	26.9	134		60	140
		Decachlorobiphenyl	1	20	25.9	130		60	140
		Tetrachloro-m-xylene	2	20	26.3	132		60	140
		Decachlorobiphenyl	2	20	28.7	144	*	60	140
P5225-03DL	3A-3B-3C-ABDL	Tetrachloro-m-xylene	1	20	0	0	*	32	144

Surrogate Summary

SDG No.: P5225

Client: ATC Group Services LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
P5225-03DL	3A-3B-3C-ABDL	Decachlorobiphenyl	1	20	0	0	*	32	175
		Tetrachloro-m-xylene	2	20	0	0	*	32	144
		Decachlorobiphenyl	2	20	0	0	*	32	175
P5225-04DL	4A-4B-4C-BASICDL	Tetrachloro-m-xylene	1	20	18.0	90		32	144
		Decachlorobiphenyl	1	20	74.0	370	*	32	175
		Tetrachloro-m-xylene	2	20	15.8	79		32	144
		Decachlorobiphenyl	2	20	78.2	391	*	32	175
I.BLK-PO108532.D	PIBLK-PO108532.D	Tetrachloro-m-xylene	1	20	26.1	131		60	140
		Decachlorobiphenyl	1	20	25.2	126		60	140
		Tetrachloro-m-xylene	2	20	26.4	132		60	140
		Decachlorobiphenyl	2	20	28.8	144	*	60	140