



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5225

Client: ATC Group Services LLC

Analytical Method: **8082A** **Datafile :** PO108510.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165544BS	AR1016	166.5	168	ug/kg	101				71	120	
	AR1260	166.5	174	ug/kg	105				65	130	