

Pratik Patel

From: Sohil Jodhani
Sent: Friday, December 27, 2024 9:54 AM
To: Deepak Parmar
Subject: Fw: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL
Attachments: cccheck.pdf; P5232-04(E28R8)B.pdf; P5155-15(E28W0)B.pdf; P5155-14(E28T9)B.pdf; P5155-12(E28T7)B.pdf; P5155-10(E28T5)B.pdf; P5155-09(E28T4)B.pdf; P5232-08(E28S2)B.pdf; P5232-06(E28S0)B.pdf; P5155-07(E28R2)B.pdf; P5155-06(E28R1)B.pdf; P5232-01(E28R5)B.pdf; P5155-05(E28R0)B.pdf

Please save in PM folder.

Thanks and Regards

Sohil Jodhani
QA/QC Director
Chemtech
An Alliance Technical Group Company
908-728-3152

From: Irving, Madison <Madison.Irving@gdit.com>
Sent: Friday, December 27, 2024 9:39:46 AM
To: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>; Mohammad Ahmed <mohammad.ahmed@alliancetg.com>
Cc: R5 - R5RSCC@epa.gov (R5RSCC@epa.gov) <R5RSCC@epa.gov>
Subject: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good morning ACE,

Regarding the laboratory's reported issue for Case 51847/SDG E28R5 & E28Q6, please see the Region and QSS input below and let me know if you have any questions.

LAB: "Lab has received soil samples for VOA analysis. Lab has performed few samples reanalysis for low level VOA analysis and closing CCV has more than two analytes outside the QC limits. Since, these were the re-analysis therefore Lab will report both original and re-analysis with closing CCV failure in final electronic deliverables.

Please see attachment for your reference."

QSS INPUT: The laboratory is expected to comply with the SOW requirements specified in Exhibit D, Sections 9.5.6 and 11.4.

REGION: "The client accepts the lab's proposal of reporting both sets of data."

***** Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division
Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: R5RSCC <R5RSCC@epa.gov>
Sent: Friday, December 27, 2024 9:17 AM
To: Danganan, Eloise B <Eloise.Danganan@gdit.com>
Cc: Irving, Madison <Madison.Irving@gdit.com>
Subject: RE: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good morning,

Sorry for the delayed response. I was waiting to hear back from the client.

The client accepts the lab's proposal of reporting both sets of data.

Thanks,

Darcie D. Tone-Pah-Hote

Regional Sample Control Coordinator

LSASD-SQAB
U.S. EPA Region 5
77 W Jackson Blvd
Chicago, IL 60604
Cubicle #: 06078
Phone #: 312-886-4298

From: Danganan, Eloise B <Eloise.Danganan@gdit.com>
Sent: Tuesday, December 24, 2024 9:48 AM
To: R5RSCC <R5RSCC@epa.gov>; Tone-Pah-Hote, Darcie <TonePahHote.Darcie@epa.gov>
Subject: RE: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good morning Region 5,

Please advise on the issue below involving Case 51847/SDGs E28R5 and E28Q6. If a response was already sent to Madison Irving, please forward it to me. She is out of the office and I do not have access to her email.

Thank you,

ELOISE DANGANAN
Environmental Analyst
GDIT Federal Civilian Division
Under contract to the EPA
T 703.254.0067
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

LEAVE ALERT: none

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Irving, Madison <Madison.Irving@gdit.com>
Sent: Friday, December 20, 2024 1:11 PM
To: R5 - R5RSCC@epa.gov (<R5RSCC@epa.gov>) <R5RSCC@epa.gov>
Subject: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

Good afternoon Region 5,

Please advise on the laboratory's reported issue below for Case 51847/SDG E28R5 & E28Q6.

LAB: "Lab has received soil samples for VOA analysis. Lab has performed few samples reanalysis for low level VOA analysis and closing CCV has more than two analytes outside the QC limits. Since, these were the re-analysis therefore Lab will report both original and re-analysis with closing CCV failure in final electronic deliverables.

Please see attachment for your reference."

QSS Input:

The laboratory is expected to comply with the SOW requirements specified in Exhibit -D-Sections, 9.5.6 and 11.4.

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division
Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: DASsupport
Sent: Friday, December 20, 2024 11:50 AM
To: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>
Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>
Subject: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

Thank you for contacting DAS Support. This is to confirm that your email has been received and is currently being addressed. An updated response will be provided as soon as it is available.

Thank you,

The DAS Support Team
DASsupport@gdit.com

GDIT Federal Civilian Division
Under contract to the EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>
Sent: Friday, December 20, 2024 11:38 AM
To: DASsupport <DASsupport@gdit.com>
Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>
Subject: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello,





Lab is sending this email with regards to case 51847.

Lab has received soil samples for VOA analysis. Lab has performed few samples reanalysis for low level VOA analysis and closing CCV has more than two analytes outside the QC limits. Since, these were the re-analysis therefore Lab will report both original and re-analysis with closing CCV failure in final electronic deliverables.

Please see attachment for your reference.

Thanks & Regards,



Sohil Jodhani
QA/QC Director
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3152
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com     

From: Maheshbhai Dadoda <Maheshbhai.Dadoda@alliancetg.com>
Sent: Friday, December 20, 2024 11:17 AM
To: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>
Subject: VD121224-SDG 5232,5155

Sequence has endcheck standard failure(more than 2 failed above 50%) and we have re-analysis samples affected due to that

All analysis are ok(no internal surrogate failed no over-range analytes) except 5155-05RE, 06RE,has internal standard failure and need medium level analysis

Lab will report all analysis as confirm analysis with failed CCC standard

Thanks

Mahesh