

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51847 MA No. : SDG No.: ME28R5
 Initial Calibration Verification Source : MP83508
 Continuing Calibration Verification Source : MP83511
 Run Batch: LB134057 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV009				ID: CCV081				ID: CCV082		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	96	1	400000	410000	102	0	400000	101	0
Calcium	10000	10000	104	0	400000	410000	103	0	400000	101	0
Iron	10000	11000	105	0	400000	420000	104	0	410000	102	0
Magnesium	6000	6300	105	2	400000	410000	102	0	400000	101	0
Potassium	9900	9200	93	3	150000	150000	103	0	150000	101	0
Sodium	10000	9100	91	4	400000	410000	102	1	410000	102	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51847 MA No. : SDG No.: ME28R5
 Initial Calibration Verification Source : MP83508
 Continuing Calibration Verification Source : MP83511
 Run Batch: LB134057 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV083				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	410000	103	0			
Calcium					400000	410000	104	0			
Iron					400000	420000	104	1			
Magnesium					400000	410000	103	0			
Potassium					150000	150000	103	0			
Sodium					400000	410000	102	0			

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51847 MA No. : SDG No.: ME28R5
 Initial Calibration Verification Source : MP83508
 Continuing Calibration Verification Source : MP83511
 Run Batch: LB134069 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	98	1	400000	400000	101	1	410000	101	1
Calcium	10000	10000	102	1	400000	410000	103	1	410000	103	1
Iron	10000	11000	106	0	400000	410000	103	1	420000	104	1
Magnesium	6000	6300	105	3	400000	410000	102	1	410000	103	1
Potassium	9900	8900	90	1	150000	150000	102	1	150000	102	1
Sodium	10000	9700	97	4	400000	400000	101	1	410000	103	0