

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51847 MA No.: _____ SDG No.: ME28R5
Instrument ID: KONELAB Start Date: 12/13/2024
Analytical Method: Spectrophotometry Run Batch: LB133955
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Cyanide	0	0.0	0	5	3.6	-28	10	8.4	-16

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51847 MA No.: SDG No.: ME28R5
Instrument ID: KONELAB Start Date: 12/13/2024
Analytical Method: Spectrophotometry Run Batch: LB133955
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Cyanide	100	100	0	250	250	0	500	500	0