

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51847 MA No.: _____ SDG No.: ME28R5
 Instrument ID: P8 Start Date: 12/13/2024
 Analytical Method: ICP-MS Run Batch: LB133950
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	0	0.0	0	2	2.1	5	50	53	6
Arsenic	0	0.0	0	1	1.1	10	50	52	4
Barium	0	0.0	0	10	11	10	250	270	8
Beryllium	0	0.0	0	1	1.1	10	50	54	8
Cadmium	0	0.0	0	1	1.1	10	50	54	8
Chromium	0	0.0	0	2	2.1	5	50	51	2
Cobalt	0	0.0	0	1	1.1	10	50	52	4
Copper	0	0.0	0	2	2.1	5	500	550	10
Lead	0	0.0	0	1	1.0	0	250	260	4
Manganese	0	0.0	0	1	1.1	10	500	510	2
Nickel	0	0.0	0	1	1.1	10	50	52	4
Selenium	0	0.0	0	5	5.6	12	50	54	8
Silver	0	0.0	0	1	1.1	10	50	55	10
Thallium	0	0.0	0	1	1.0	0	50	52	4
Vanadium	0	0.0	0	5	5.2	4	50	51	2
Zinc	0	0.0	0	5	5.3	6	500	520	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51847 MA No.: _____ SDG No.: ME28R5
 Instrument ID: P8 Start Date: 12/13/2024
 Analytical Method: ICP-MS Run Batch: LB133950
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	125	140	12	250	260	4	500	510	2
Arsenic	125	130	4	250	260	4	500	510	2
Barium	625	660	6	1250	1300	4	2500	2500	0
Beryllium	125	130	4	250	260	4	500	510	2
Cadmium	125	130	4	250	260	4	500	520	4
Chromium	125	130	4	250	250	0	500	520	4
Cobalt	125	130	4	250	270	8	500	510	2
Copper	1250	1300	4	2500	2700	8	5000	5100	2
Lead	625	650	4	1250	1300	4	2500	2500	0
Manganese	1250	1300	4	2500	2600	4	5000	5100	2
Nickel	125	130	4	250	260	4	500	490	-2
Selenium	125	130	4	250	260	4	500	510	2
Silver	125	140	12	250	260	4	500	520	4
Thallium	125	130	4	250	250	0	500	500	0
Vanadium	125	120	-4	250	260	4	500	510	2
Zinc	1250	1300	4	2500	2700	8	5000	5200	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51847 MA No.: SDG No.: ME28R5
Instrument ID: P8 Start Date: 12/13/2024
Analytical Method: ICP-MS Run Batch: LB133950
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	1000	990	-1						
Arsenic	1000	990	-1						
Barium	5000	5000	0						
Beryllium	1000	990	-1						
Cadmium	1000	990	-1						
Chromium	1000	990	-1						
Cobalt	1000	990	-1						
Copper	10000	9900	-1						
Lead	5000	5000	0						
Manganese	10000	9900	-1						
Nickel	1000	1000	0						
Selenium	1000	990	-1						
Silver	1000	980	-2						
Thallium	1000	1000	0						
Vanadium	1000	990	-1						
Zinc	10000	9900	-1						