

Pratik Patel

From: Danganan, Eloise B <Eloise.Danganan@gdit.com>
Sent: Thursday, December 12, 2024 1:38 PM
To: Sohil Jodhani; Mohammad Ahmed; Nimisha Pandya
Cc: REGION_3 (burman.jarmael@epa.gov)
Subject: REGION 3 | CASE 51901 | LAB ACE | SDG C0AD3 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL
Attachments: PR069445_D1a.pdf; PR069448_D1a.pdf; quantPR069445.pdf; quantPR069448.pdf; 2C.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon Sohil,

Please see the Region's response below regarding the laboratory's reported issue for Case 51901/SDG C0AD3. Please include detailed information about this issue in the SDG Narrative and let me know if you have any questions.

LAB: "Lab has received soil samples for PCB analysis. Lab has performed the undiluted analysis for the sample C0AD3 & C0AQ8 and sample found positive with extremely high concentration of PCB analytes. Due to very high concentrations of target analytes, surrogates are affected and outside the QC limits as you can see attached form-1, quant reports and surrogate summary for your reference. In this case, Lab will report undiluted PCB analysis with surrogate failure for these samples and further dilution in final electronic deliverables. Please see attached for your reference."

REGION: "Inform ACE the Region is in agreement with their below written approach; have ACE make note of the issue in their SDG Narrative and proceed with the analysis of the samples."

***** Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

ELOISE DANGANAN
Environmental Analyst
GDIT Federal Civilian Division
Under contract to the EPA
T 703.254.0067
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

LEAVE ALERT:

- December 19-23

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Burman, Jarmael <Burman.Jarmael@epa.gov>

Sent: Thursday, December 12, 2024 1:36 PM

To: Danganan, Eloise B <Eloise.Danganan@gdit.com>

Subject: RE: REGION 3 | CASE 51901 | LAB ACE | SDG C0AD3 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good afternoon Eloise,

Inform ACE the Region is in agreement with their below written approach; have ACE make note of the issue in their SDG Narrative and proceed with the analysis of the samples.

Jarmael Burman

US EPA Region 3 - CLP RR/RSCC/DDS/QA Chemist/DAS PO/EEOC

701 Mapes Road

Fort Meade, Maryland 20755-5350

(410) 305-2743 (office)

(410) 305-3095 (fax)

From: Danganan, Eloise B <Eloise.Danganan@gdit.com>

Sent: Thursday, December 12, 2024 1:31 PM

To: Burman, Jarmael <Burman.Jarmael@epa.gov>

Subject: REGION 3 | CASE 51901 | LAB ACE | SDG C0AD3 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good afternoon Jay,

Please advise on the issue reported by ACE below for Case 51901/SDG C0AD3.

LAB: Lab has received soil samples for PCB analysis. Lab has performed the undiluted analysis for the sample C0AD3 & C0AQ8 and sample found positive with extremely high concentration of PCB analytes. Due to very high concentrations of target analytes, surrogates are affected and outside the QC limits as you can see attached form-1, quant reports and surrogate summary for your reference. In this case, Lab will report undiluted PCB analysis with surrogate failure for these samples and further dilution in final electronic deliverables.

Please see attached for your reference.

Thank you,

ELOISE DANGANAN
Environmental Analyst
GDIT Federal Civilian Division
Under contract to the EPA
T 703.254.0067
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

LEAVE ALERT:

- December 19-23

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>

Sent: Thursday, December 12, 2024 12:23 PM

To: DASsupport <DASsupport@gdit.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>

Subject: REGION 3 | CASE 51901 | LAB ACE | SDG C0AD3 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello,

Lab is sending this email with regards to case 51901 & SDG C0AD3.

Lab has received soil samples for PCB analysis. Lab has performed the undiluted analysis for the sample C0AD3 & C0AQ8 and sample found positive with extremely high concentration of PCB analytes. Due to very high concentrations of target analytes, surrogates are affected and outside the QC limits as you can see attached form-1, quant reports and surrogate summary for your reference. In this case, Lab will report undiluted PCB analysis with surrogate failure for these samples and further dilution in final electronic deliverables.

Please see attached for your reference.

Thanks & Regards,



Sohil Jodhani

QA/QC Director

An Alliance Technical Group Company

Main: 908-789-8900

Direct: 908-728-3152

Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com

