

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: P5236

MATRIX: Solid

METHOD: 9045D,9056A,9060A

	NA	NO	YES
1. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
2. Matrix Spike Duplicate Recoveries Met Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Blank Spike met requirements for all samples. The Matrix Spike (TAPIAL3-SB03D-3-120724-00-T01MS) analysis met criteria for all samples except for Fluoride due to sample matrix interference. The Matrix Spike Duplicate (TAPIAL3-SB03D-3-120724-00-T01MSD) analysis met criteria for all samples except for Fluoride due to sample matrix interference.			
3. Sample Duplicate Analysis Met QC Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
4. Digestion Holding Time Met		✓	
If not met, list number of days exceeded for each sample:			
The Holding Times were met for all samples except for TAPIAL3-SB03D-3-120724-00-T01 of pH as sample received out of holding time.			

ADDITIONAL COMMENTS:

QA REVIEW

Date