

**ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092**

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

**METALS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: P5236

MATRIX: Solid

METHOD: 6020B,7471B

|  | NA | NO | YES |
|--|----|----|-----|
| 1. Calibration Summary met criteria.   |    |    | ✓   |
| 2. ICP Interference Check Sample Results Summary Submitted.  |    |    | ✓   |
| 3. Serial Dilution Summary (if applicable) Submitted.  |    | ✓  |     |
| The Serial Dilution met criteria for all samples.  |    |    |     |
| 4. Laboratory Control Sample Summary (if applicable) Submitted.  |    |    | ✓   |
| 5. Blank Contamination - If yes, list compounds and concentrations in each blank:  |    | ✓  |     |
| 6. Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria   |    | ✓  |     |
| If not met, list those compounds and their recoveries which fall outside the acceptable range.   |    |    |     |
| The Matrix Spike (TAPIAL3-SB04I-10-120324-00-T1MS) analysis met criteria for all samples except for Antimony, Arsenic, Beryllium, Cadmium, Chromium, Cobalt, Nickel, Selenium, Silver, Thallium, Vanadium due to matrix interference. The Matrix Spike Duplicate (TAPIAL3-SB04I-10-120324-00-T1MSD) analysis met criteria for all samples except for Antimony, Arsenic, Beryllium, Cadmium, Chromium, Cobalt, Nickel, Selenium, Silver, Thallium, Vanadium due to matrix interference. |    |    |     |
| 7. Sample Duplicate Analysis Met QC Criteria   |    |    | ✓   |
| If not met, list those compounds and their recoveries which fall outside the acceptable range.   |    |    |     |
| 8. Digestion Holding Time Met  |    |    | ✓   |
| If not met, list number of days exceeded for each sample:  |    |    |     |
| 9. Analysis Holding Time Met   |    |    | ✓   |
| If not met, list those compounds and their recoveries which fall outside the acceptable range.   |    |    |     |

ADDITIONAL COMMENTS: Sample P5236-01 reported with straight 5X dilution due to high interferent samples.

Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.

---

QA REVIEW

---

Date