

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P5236

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133872

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10	10	100	90-110	11/18/2024
Chloride	mg/L	3	3	100	90-110	11/18/2024
Fluoride	mg/L	2	2	100	90-110	11/18/2024
Nitrite	mg/L	3	3	100	90-110	11/18/2024
Nitrate	mg/L	2.5	2.5	100	90-110	11/18/2024
Sulfate	mg/L	14.6	15	97	90-110	11/18/2024
Orthophosphate as P	mg/L	5	5	100	90-110	11/18/2024
Sample ID: CCV1						
Bromide	mg/L	10.7	10	107	90-110	12/11/2024
Chloride	mg/L	3.2	3	107	90-110	12/11/2024
Fluoride	mg/L	2.1	2	105	90-110	12/11/2024
Nitrite	mg/L	3.2	3	107	90-110	12/11/2024
Nitrate	mg/L	2.7	2.5	108	90-110	12/11/2024
Sulfate	mg/L	15.4	15	103	90-110	12/11/2024
Orthophosphate as P	mg/L	5.3	5	106	90-110	12/11/2024
Sample ID: CCV2						
Bromide	mg/L	10.7	10	107	90-110	12/11/2024
Chloride	mg/L	3.2	3	107	90-110	12/11/2024
Fluoride	mg/L	2.1	2	105	90-110	12/11/2024
Nitrite	mg/L	3.2	3	107	90-110	12/11/2024
Nitrate	mg/L	2.7	2.5	108	90-110	12/11/2024
Sulfate	mg/L	15.3	15	102	90-110	12/11/2024
Orthophosphate as P	mg/L	5.3	5	106	90-110	12/11/2024

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P5236

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133903

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.00	7	100	90-110	12/12/2024
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	12/12/2024
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	12/12/2024

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P5236

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133969

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
TOC		mg/L	934	1000	93	90-110	11/12/2024
Sample ID:	CCV1						
TOC		mg/L	986	1000	99	90-110	12/17/2024
Sample ID:	CCV2						
TOC		mg/L	1060	1000	106	90-110	12/17/2024