

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions

SDG No.: P5236

Contract: WEST04

Lab Code: CHEM

Case No.: P5236

SAS No.: P5236

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165560BS Mercury	mg/Kg	0.26	0.23		89	80 - 124	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions **SDG No.:** P5236
Contract: WEST04 **Lab Code:** CHEM **Case No.:** P5236 **SAS No.:** P5236

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165717BS							
Aluminum	mg/Kg	710	707		100	78 - 124	P
Antimony	mg/Kg	35.7	37.4		105	72 - 124	P
Arsenic	mg/Kg	35.7	36.7		103	82 - 118	P
Barium	mg/Kg	180	185		103	86 - 116	P
Beryllium	mg/Kg	35.7	35.5		99	80 - 120	P
Cadmium	mg/Kg	35.7	37.5		105	84 - 116	P
Calcium	mg/Kg	3600	3700		103	86 - 118	P
Chromium	mg/Kg	35.7	36.7		103	83 - 119	P
Cobalt	mg/Kg	35.7	36.4		102	84 - 115	P
Copper	mg/Kg	360	367		102	84 - 119	P
Iron	mg/Kg	1800	1870		104	81 - 124	P
Lead	mg/Kg	180	182		101	84 - 118	P
Magnesium	mg/Kg	3600	3570		99	80 - 123	P
Manganese	mg/Kg	360	362		101	85 - 116	P
Nickel	mg/Kg	35.7	35.6		100	84 - 119	P
Potassium	mg/Kg	1800	1790		99	85 - 119	P
Selenium	mg/Kg	35.7	37.5		105	80 - 119	P
Silver	mg/Kg	35.7	38.4		108	83 - 118	P
Sodium	mg/Kg	3600	3620		101	79 - 125	P
Thallium	mg/Kg	35.7	36.6		102	83 - 118	P
Vanadium	mg/Kg	35.7	36.2		101	82 - 116	P
Zinc	mg/Kg	360	366		102	82 - 119	P