

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ATCE02

Lab Code: CHEM Case No.: P5240 SAS No.: P5240 SDG No.: P5240

Instrument ID: MSVOA_N Calibration Date/Time: 12/10/2024 10:17

Lab File ID: VN085155.D Init. Calib. Date(s): 11/22/2024 11/22/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:22 13:45

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.445	1.502		4.02	20
Toluene	0.855	0.947		10.76	20
Ethyl Benzene	1.840	2.077		12.88	20
m/p-Xylenes	0.700	0.799		14.14	20
o-Xylene	0.666	0.751		12.76	20
1,2-Dichloroethane-d4	0.712	0.624		-12.36	20
Dibromofluoromethane	0.337	0.318		-5.64	20
Toluene-d8	1.208	1.112		-7.95	20
4-Bromofluorobenzene	0.452	0.429		-5.09	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.