

Surrogate Summary

SDG No.: **P5240**

Client: **ATC Group Services LLC**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5240-01	OF-1	1,2-Dichloroethane-d4	50	54.2	108	74	125
		Dibromofluoromethane	50	50.1	100	75	124
		Toluene-d8	50	49.4	99	86	113
		4-Bromofluorobenzene	50	43.2	86	77	121
P5240-04	OF-2	1,2-Dichloroethane-d4	50	55.0	110	74	125
		Dibromofluoromethane	50	50.0	100	75	124
		Toluene-d8	50	47.4	95	86	113
		4-Bromofluorobenzene	50	42.6	85	77	121
P5240-05	OF-3	1,2-Dichloroethane-d4	50	54.5	109	74	125
		Dibromofluoromethane	50	51.2	102	75	124
		Toluene-d8	50	48.3	97	86	113
		4-Bromofluorobenzene	50	41.5	83	77	121
VN1210WBL01	VN1210WBL01	1,2-Dichloroethane-d4	50	53.0	106	74	125
		Dibromofluoromethane	50	51.7	103	75	124
		Toluene-d8	50	49.3	99	86	113
		4-Bromofluorobenzene	50	43.9	88	77	121
VN1210WBS01	VN1210WBS01	1,2-Dichloroethane-d4	50	49.5	99	74	125
		Dibromofluoromethane	50	51.8	103	75	124
		Toluene-d8	50	50.5	101	86	113
		4-Bromofluorobenzene	50	50.3	101	77	121
VN1210WBSD01	VN1210WBSD01	1,2-Dichloroethane-d4	50	50.3	101	74	125
		Dibromofluoromethane	50	52.0	104	75	124
		Toluene-d8	50	49.6	99	86	113
		4-Bromofluorobenzene	50	50.9	102	77	121