

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5240
Client: ATC Group Services LLC
Analytical Method: SW8260-Low **Datafile :** VN085158.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1210WBS01	Benzene	20	18.0	ug/L	90			82	109	
	Toluene	20	18.7	ug/L	94			82	110	
	Ethyl Benzene	20	18.4	ug/L	92			83	109	
	m/p-Xylenes	40	37.8	ug/L	95			82	110	
	o-Xylene	20	19.2	ug/L	96			83	109	