

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5240

Client: ATC Group Services LLC

Analytical Method: SW8260-Low

Datafile : VN085159.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1210WBSD01	Benzene	20	17.7	ug/L	89	1		82	109	20
	Toluene	20	18.9	ug/L	95	1		82	110	20
	Ethyl Benzene	20	18.1	ug/L	91	1		83	109	20
	m/p-Xylenes	40	37.1	ug/L	93	2		82	110	20
	o-Xylene	20	18.4	ug/L	92	4		83	109	20