

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ATCE02
 Lab Code: CHEM Case No.: P5240 SAS No.: P5240 SDG NO.: P5240
 Lab File ID: VN085155.D Date Analyzed: 12/10/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:17
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	161771	8.22	254585	9.10	227639	11.87
UPPER LIMIT	323542	8.724	509170	9.6	455278	12.365
LOWER LIMIT	80885.5	7.724	127293	8.6	113820	11.365
EPA SAMPLE NO.						
OF-1	129349	8.22	237566	9.10	204448	11.87
OF-2	134322	8.22	251295	9.10	208512	11.87
OF-3	130450	8.22	239679	9.10	197437	11.87
VN1210WBL01	144184	8.22	251448	9.10	221103	11.87
VN1210WBS01	148284	8.22	241760	9.10	211196	11.87
VN1210WBSD01	142566	8.22	233684	9.10	205174	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.