

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ATCE02
 Lab Code: CHEM Case No.: P5240 SAS No.: P5240 SDG NO.: P5240
 Lab File ID: VN085155.D Date Analyzed: 12/10/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:17
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	114435	13.788				
UPPER LIMIT	228870	14.288				
LOWER LIMIT	57217.5	13.288				
EPA SAMPLE NO.						
OF-1	82722	13.79				
OF-2	88019	13.79				
OF-3	75083	13.79				
VN1210WBL01	86343	13.79				
VN1210WBS01	107293	13.79				
VN1210WBSD01	102534	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.