



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Weston Solutions SDG No.: P5241
Contract: WEST04 Lab Code: CHEM Case No.: P5241 SAS No.: P5241
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Aluminum	20.2	20.0	101	70 - 130	P	12/11/2024	18:45	LB133901
	Antimony	2.14	2.0	107	70 - 130	P	12/11/2024	18:45	LB133901
	Arsenic	1.16	1.0	116	70 - 130	P	12/11/2024	18:45	LB133901
	Barium	10.5	10.0	105	70 - 130	P	12/11/2024	18:45	LB133901
	Beryllium	1.10	1.0	110	70 - 130	P	12/11/2024	18:45	LB133901
	Cadmium	1.06	1.0	106	70 - 130	P	12/11/2024	18:45	LB133901
	Calcium	525	500	105	70 - 130	P	12/11/2024	18:45	LB133901
	Chromium	2.16	2.0	108	70 - 130	P	12/11/2024	18:45	LB133901
	Cobalt	1.10	1.0	110	50 - 150	P	12/11/2024	18:45	LB133901
	Copper	2.02	2.0	101	70 - 130	P	12/11/2024	18:45	LB133901
	Iron	55.2	50.0	110	70 - 130	P	12/11/2024	18:45	LB133901
	Lead	0.99	1.0	99	70 - 130	P	12/11/2024	18:45	LB133901
	Magnesium	568	500	114	70 - 130	P	12/11/2024	18:45	LB133901
	Manganese	1.12	1.0	112	50 - 150	P	12/11/2024	18:45	LB133901
	Nickel	1.08	1.0	108	70 - 130	P	12/11/2024	18:45	LB133901
	Potassium	519	500	104	70 - 130	P	12/11/2024	18:45	LB133901
	Selenium	5.54	5.0	111	70 - 130	P	12/11/2024	18:45	LB133901
	Silver	1.06	1.0	106	70 - 130	P	12/11/2024	18:45	LB133901
	Sodium	553	500	111	70 - 130	P	12/11/2024	18:45	LB133901
	Thallium	0.98	1.0	98	70 - 130	P	12/11/2024	18:45	LB133901
	Vanadium	5.19	5.0	104	70 - 130	P	12/11/2024	18:45	LB133901
	Zinc	5.93	5.0	119	50 - 150	P	12/11/2024	18:45	LB133901