



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions SDG No.: P5241
Contract: WEST04 Lab Code: CHEM Case No.: P5241 SAS No.: P5241

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	3.62	+/-20.0	J	10.0	20.0	P	12/11/2024	18:02	LB133901
	Antimony	2.00	+/-2.00	U	0.25	2.00	P	12/11/2024	18:02	LB133901
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	18:02	LB133901
	Barium	10.0	+/-10.0	U	1.25	10.0	P	12/11/2024	18:02	LB133901
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	18:02	LB133901
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/11/2024	18:02	LB133901
	Calcium	500	+/-500	U	190	500	P	12/11/2024	18:02	LB133901
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/11/2024	18:02	LB133901
	Cobalt	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	18:02	LB133901
	Copper	2.00	+/-2.00	U	1.50	2.00	P	12/11/2024	18:02	LB133901
	Iron	50.0	+/-50.0	U	25.0	50.0	P	12/11/2024	18:02	LB133901
	Lead	1.00	+/-1.00	U	0.75	1.00	P	12/11/2024	18:02	LB133901
	Magnesium	500	+/-500	U	190	500	P	12/11/2024	18:02	LB133901
	Manganese	1.00	+/-1.00	U	0.75	1.00	P	12/11/2024	18:02	LB133901
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	18:02	LB133901
	Potassium	500	+/-500	U	190	500	P	12/11/2024	18:02	LB133901
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/11/2024	18:02	LB133901
	Silver	1.00	+/-1.00	U	0.50	1.00	P	12/11/2024	18:02	LB133901
	Sodium	500	+/-500	U	190	500	P	12/11/2024	18:02	LB133901
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	12/11/2024	18:02	LB133901
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/11/2024	18:02	LB133901
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/11/2024	18:02	LB133901

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>P5241</u>								
Contract: <u>WEST04</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5241</u>	SAS No.: <u>P5241</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	20.0	+/-20.0	U	10.0	20.0	P	12/11/2024	18:15	LB133901
	Antimony	2.00	+/-2.00	U	0.25	2.00	P	12/11/2024	18:15	LB133901
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	18:15	LB133901
	Barium	10.0	+/-10.0	U	1.25	10.0	P	12/11/2024	18:15	LB133901
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	18:15	LB133901
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/11/2024	18:15	LB133901
	Calcium	500	+/-500	U	190	500	P	12/11/2024	18:15	LB133901
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/11/2024	18:15	LB133901
	Cobalt	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	18:15	LB133901
	Copper	0.40	+/-2.00	J	1.50	2.00	P	12/11/2024	18:15	LB133901
	Iron	50.0	+/-50.0	U	25.0	50.0	P	12/11/2024	18:15	LB133901
	Lead	0.12	+/-1.00	J	0.75	1.00	P	12/11/2024	18:15	LB133901
	Magnesium	500	+/-500	U	190	500	P	12/11/2024	18:15	LB133901
	Manganese	1.00	+/-1.00	U	0.75	1.00	P	12/11/2024	18:15	LB133901
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	18:15	LB133901
	Potassium	53.1	+/-500	J	190	500	P	12/11/2024	18:15	LB133901
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/11/2024	18:15	LB133901
	Silver	1.00	+/-1.00	U	0.50	1.00	P	12/11/2024	18:15	LB133901
	Sodium	103	+/-500	J	190	500	P	12/11/2024	18:15	LB133901
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	12/11/2024	18:15	LB133901
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/11/2024	18:15	LB133901
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/11/2024	18:15	LB133901
CCB02	Aluminum	20.0	+/-20.0	U	10.0	20.0	P	12/11/2024	19:21	LB133901
	Antimony	2.00	+/-2.00	U	0.25	2.00	P	12/11/2024	19:21	LB133901
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	19:21	LB133901
	Barium	10.0	+/-10.0	U	1.25	10.0	P	12/11/2024	19:21	LB133901
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	19:21	LB133901
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/11/2024	19:21	LB133901
	Calcium	500	+/-500	U	190	500	P	12/11/2024	19:21	LB133901
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/11/2024	19:21	LB133901
	Cobalt	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	19:21	LB133901
	Copper	2.00	+/-2.00	U	1.50	2.00	P	12/11/2024	19:21	LB133901
	Iron	50.0	+/-50.0	U	25.0	50.0	P	12/11/2024	19:21	LB133901
	Lead	1.00	+/-1.00	U	0.75	1.00	P	12/11/2024	19:21	LB133901
	Magnesium	500	+/-500	U	190	500	P	12/11/2024	19:21	LB133901
	Manganese	1.00	+/-1.00	U	0.75	1.00	P	12/11/2024	19:21	LB133901
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	19:21	LB133901
	Potassium	500	+/-500	U	190	500	P	12/11/2024	19:21	LB133901
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/11/2024	19:21	LB133901

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions **SDG No.:** P5241
Contract: WEST04 **Lab Code:** CHEM **Case No.:** P5241 **SAS No.:** P5241

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.00	+/-1.00	U	0.50	1.00	P	12/11/2024	19:21	LB133901
	Sodium	500	+/-500	U	190	500	P	12/11/2024	19:21	LB133901
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	12/11/2024	19:21	LB133901
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/11/2024	19:21	LB133901
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/11/2024	19:21	LB133901
CCB03	Aluminum	20.0	+/-20.0	U	10.0	20.0	P	12/11/2024	19:55	LB133901
	Antimony	0.12	+/-2.00	J	0.25	2.00	P	12/11/2024	19:55	LB133901
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	19:55	LB133901
	Barium	10.0	+/-10.0	U	1.25	10.0	P	12/11/2024	19:55	LB133901
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	19:55	LB133901
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	12/11/2024	19:55	LB133901
	Calcium	500	+/-500	U	190	500	P	12/11/2024	19:55	LB133901
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	12/11/2024	19:55	LB133901
	Cobalt	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	19:55	LB133901
	Copper	2.00	+/-2.00	U	1.50	2.00	P	12/11/2024	19:55	LB133901
	Iron	50.0	+/-50.0	U	25.0	50.0	P	12/11/2024	19:55	LB133901
	Lead	0.13	+/-1.00	J	0.75	1.00	P	12/11/2024	19:55	LB133901
	Magnesium	500	+/-500	U	190	500	P	12/11/2024	19:55	LB133901
	Manganese	1.00	+/-1.00	U	0.75	1.00	P	12/11/2024	19:55	LB133901
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	12/11/2024	19:55	LB133901
	Potassium	500	+/-500	U	190	500	P	12/11/2024	19:55	LB133901
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	12/11/2024	19:55	LB133901
	Silver	1.00	+/-1.00	U	0.50	1.00	P	12/11/2024	19:55	LB133901
	Sodium	96.3	+/-500	J	190	500	P	12/11/2024	19:55	LB133901
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	12/11/2024	19:55	LB133901
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	12/11/2024	19:55	LB133901
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	12/11/2024	19:55	LB133901