

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions
Lab Code: CHEM Case no.: P5241
Instrument ID: P8
Run Number: LB133901

Contract: WEST04
Sas No.: P5241 SDG No.: P5241
Start Date : 12/11/2024
End Date : 12/11/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1639	100		100							
S2	S2	1645	104		105							
S3	S3	1648	104		105							
S4	S4	1651	105		105							
S5	S5	1654	103		102							
S6	S6	1657	105		101							
S7	S7	1700	105		99							
S8	S8	1703	90		80							
ICV01	ICV01	1725	106		105							
LLICV	LLICV	1759	106		105							
ICB01	ICB01	1802	104		104							
ICSA01	ICSA01	1805	102		95							
ICSAB01	ICSAB01	1808	105		97							
CCV01	CCV01	1811	103		94							
CCB01	CCB01	1815	109		109							
CRI	CRI	1845	107		106							
ZZZZZZ	ZZZZZZ	1848										
ZZZZZZ	ZZZZZZ	1852										
ZZZZZZ	ZZZZZZ	1854										
ZZZZZZ	ZZZZZZ	1858										
ZZZZZZ	ZZZZZZ	1901										
ZZZZZZ	ZZZZZZ	1904										
ZZZZZZ	ZZZZZZ	1907										
ZZZZZZ	ZZZZZZ	1910										
ZZZZZZ	ZZZZZZ	1913										
CCV02	CCV02	1916	100		92							
CCB02	CCB02	1921	107		108							
PB165562BL	PB165562BL	1924	104		105							
PB165562BS	PB165562BS	1927	105		101							
P5241-01	TAPIAL2-SB02	1930	108		108							
P5241-01DUP	TAPIAL2-SB02	1933	108		107							
P5241-01L	TAPIAL2-SB02	1937	109		108							
P5241-01MS	TAPIAL2-SB02	1940	109		107							
P5241-01MSD	TAPIAL2-SB02	1943	108		107							

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions

Contract: WEST04

Lab Code: CHEM Case no.: P5241

Sas No.: P5241 SDG No.: P5241

Instrument ID: P8

Start Date : 12/11/2024

Run Number: LB133901

End Date : 12/11/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
P5241-01A	TAPIAL2-SB02	1949	100		100							
CCV03	CCV03	1952	101		91							
CCB03	CCB03	1955	106		107							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions
Lab Code: CHEM Case no.: P5241
Instrument ID: P8
Run Number: LB133901

Contract: WEST04
Sas No.: P5241 SDG No.: P5241
Start Date : 12/11/2024
End Date : 12/11/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1639	100									
S2	S2	1645	95									
S3	S3	1648	96									
S4	S4	1651	93									
S5	S5	1654	92									
S6	S6	1657	92									
S7	S7	1700	89									
S8	S8	1703	73									
ICV01	ICV01	1725	99									
LLICV	LLICV	1759	99									
ICB01	ICB01	1802	98									
ICSA01	ICSA01	1805	87									
ICSAB01	ICSAB01	1808	88									
CCV01	CCV01	1811	83									
CCB01	CCB01	1815	100									
CRI	CRI	1845	98									
ZZZZZZ	ZZZZZZ	1848										
ZZZZZZ	ZZZZZZ	1852										
ZZZZZZ	ZZZZZZ	1854										
ZZZZZZ	ZZZZZZ	1858										
ZZZZZZ	ZZZZZZ	1901										
ZZZZZZ	ZZZZZZ	1904										
ZZZZZZ	ZZZZZZ	1907										
ZZZZZZ	ZZZZZZ	1910										
ZZZZZZ	ZZZZZZ	1913										
CCV02	CCV02	1916	83									
CCB02	CCB02	1921	98									
PB165562BL	PB165562BL	1924	96									
PB165562BS	PB165562BS	1927	92									
P5241-01	TAPIAL2-SB02	1930	97									
P5241-01DUP	TAPIAL2-SB02	1933	98									
P5241-01L	TAPIAL2-SB02	1937	99									
P5241-01MS	TAPIAL2-SB02	1940	96									
P5241-01MSD	TAPIAL2-SB02	1943	95									

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions

Contract: WEST04

Lab Code: CHEM Case no.: P5241

Sas No.: P5241 SDG No.: P5241

Instrument ID: P8

Start Date : 12/11/2024

Run Number: LB133901

End Date : 12/11/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
P5241-01A	TAPIAL2-SB02	1949	96									
CCV03	CCV03	1952	83									
CCB03	CCB03	1955	97									