

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Weston Solutions

Contract: WEST04

Lab Code: CHEM Case no.: P5242

Sas No.: P5242 SDG No.: P5242

Instrument ID: P8

Start Date : 12/11/2024

Run Number: LB133901

End Date : 12/11/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
S0	S0	1639	100		100		100		100		100	
S2	S2	1645	104		103		104		104		104	
S3	S3	1648	104		104		104		103		106	
S4	S4	1651	103		100		103		100		106	
S5	S5	1654	99		97		100		96		103	
S6	S6	1657	99		98		102		96		105	
S7	S7	1700	92		99		101		93		104	
S8	S8	1703	80		82		86		75		89	
ICV01	ICV01	1725	100		105		107		106		107	
LLICV	LLICV	1759	105		107		105		105		105	
ICB01	ICB01	1802	103		105		105		104		105	
ICSA01	ICSA01	1805	90		95		98		90		101	
ICSAB01	ICSAB01	1808	91		101		104		95		106	
CCV01	CCV01	1811	86		98		100		89		105	
CCB01	CCB01	1815	99		105		108		107		107	
CRI	CRI	1845	101		106		108		107		106	
PB165537BL	PB165537BL	1848	101		107		108		107		106	
PB165537BS	PB165537BS	1852	94		103		104		98		107	
P5242-01	TAPIAL1-SB0	1854	96		106		109		104		107	
P5242-02	TAPIAL1-SB0	1858	96		106		108		105		108	
P5242-02DUP	TAPIAL1-SB0	1901	95		105		107		104		108	
P5242-02L	TAPIAL1-SB0	1904	95		104		106		105		107	
P5242-02MS	TAPIAL1-SB0	1907	94		105		107		102		107	
P5242-02MSD	TAPIAL1-SB0	1910	95		105		107		102		108	
P5242-02A	TAPIAL1-SB0	1913	92		102		104		101		104	
CCV02	CCV02	1916	87		94		96		86		100	
CCB02	CCB02	1921	97		103		106		105		107	

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Weston Solutions

Contract: WEST04

Lab Code: CHEM Case no.: P5242

Sas No.: P5242 SDG No.: P5242

Instrument ID: P8

Start Date : 12/11/2024

Run Number: LB133901

End Date : 12/11/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
S0	S0	1639	100		100		100		100		100	
S2	S2	1645	95		95		95		96		96	
S3	S3	1648	94		95		96		98		98	
S4	S4	1651	90		92		93		97		96	
S5	S5	1654	89		92		91		96		95	
S6	S6	1657	89		92		89		96		96	
S7	S7	1700	89		92		87		96		95	
S8	S8	1703	77		81		71		84		84	
ICV01	ICV01	1725	99		100		100		101		100	
LLICV	LLICV	1759	100		101		100		100		98	
ICB01	ICB01	1802	99		100		99		99		99	
ICSA01	ICSA01	1805	90		93		87		96		95	
ICSAB01	ICSAB01	1808	92		95		89		96		97	
CCV01	CCV01	1811	89		91		82		93		94	
CCB01	CCB01	1815	97		100		101		100		101	
CRI	CRI	1845	100		100		100		99		99	
PB165537BL	PB165537BL	1848	100		100		100		99		99	
PB165537BS	PB165537BS	1852	92		96		91		97		98	
P5242-01	TAPIAL1-SB0	1854	96		99		97		99		99	
P5242-02	TAPIAL1-SB0	1858	88		91		89		91		90	
P5242-02DUP	TAPIAL1-SB0	1901	98		101		100		102		102	
P5242-02L	TAPIAL1-SB0	1904	96		99		100		100		101	
P5242-02MS	TAPIAL1-SB0	1907	94		97		95		99		97	
P5242-02MSD	TAPIAL1-SB0	1910	93		95		95		98		98	
P5242-02A	TAPIAL1-SB0	1913	94		95		95		98		97	
CCV02	CCV02	1916	86		88		80		92		92	
CCB02	CCB02	1921	95		97		98		100		99	