

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions
Lab Code: CHEM Case no.: P5242
Instrument ID: P8
Run Number: LB133901

Contract: WEST04
Sas No.: P5242 SDG No.: P5242
Start Date : 12/11/2024
End Date : 12/11/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1639	100		100							
S2	S2	1645	104		105							
S3	S3	1648	104		105							
S4	S4	1651	105		105							
S5	S5	1654	103		102							
S6	S6	1657	105		101							
S7	S7	1700	105		99							
S8	S8	1703	90		80							
ICV01	ICV01	1725	106		105							
LLICV	LLICV	1759	106		105							
ICB01	ICB01	1802	104		104							
ICSA01	ICSA01	1805	102		95							
ICSAB01	ICSAB01	1808	105		97							
CCV01	CCV01	1811	103		94							
CCB01	CCB01	1815	109		109							
CRI	CRI	1845	107		106							
PB165537BL	PB165537BL	1848	106		106							
PB165537BS	PB165537BS	1852	107		101							
P5242-01	TAPIAL1-SB04	1854	108		107							
P5242-02	TAPIAL1-SB04	1858	109		107							
P5242-02DUP	TAPIAL1-SB04	1901	108		107							
P5242-02L	TAPIAL1-SB04	1904	107		108							
P5242-02MS	TAPIAL1-SB04	1907	108		106							
P5242-02MSD	TAPIAL1-SB04	1910	108		105							
P5242-02A	TAPIAL1-SB04	1913	105		104							
CCV02	CCV02	1916	100		92							
CCB02	CCB02	1921	107		108							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions
Lab Code: CHEM Case no.: P5242
Instrument ID: P8
Run Number: LB133901

Contract: WEST04
Sas No.: P5242 SDG No.: P5242
Start Date : 12/11/2024
End Date : 12/11/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1639	100									
S2	S2	1645	95									
S3	S3	1648	96									
S4	S4	1651	93									
S5	S5	1654	92									
S6	S6	1657	92									
S7	S7	1700	89									
S8	S8	1703	73									
ICV01	ICV01	1725	99									
LLICV	LLICV	1759	99									
ICB01	ICB01	1802	98									
ICSA01	ICSA01	1805	87									
ICSAB01	ICSAB01	1808	88									
CCV01	CCV01	1811	83									
CCB01	CCB01	1815	100									
CRI	CRI	1845	98									
PB165537BL	PB165537BL	1848	99									
PB165537BS	PB165537BS	1852	91									
P5242-01	TAPIAL1-SB04	1854	96									
P5242-02	TAPIAL1-SB04	1858	89									
P5242-02DUP	TAPIAL1-SB04	1901	100									
P5242-02L	TAPIAL1-SB04	1904	100									
P5242-02MS	TAPIAL1-SB04	1907	96									
P5242-02MSD	TAPIAL1-SB04	1910	94									
P5242-02A	TAPIAL1-SB04	1913	96									
CCV02	CCV02	1916	83									
CCB02	CCB02	1921	98									