

## Duplicate Sample Summary

|                   |                                      |                                         |          |
|-------------------|--------------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | PARSONS Main of New York, Inc.       | <b>SDG No.:</b>                         | P5246    |
| <b>Project:</b>   | Con Edison Non-MGP - Atlantic Avenue | <b>Sample ID:</b>                       | P5246-01 |
| <b>Client ID:</b> | MW7-20241209DUP                      | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|---------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| TDS     | mg/L  | +/-5                | 870              |                    | 905                 |                    | 1                  | 3.94       |      | 12/10/2024       |

## Duplicate Sample Summary

|                                                      |                                           |
|------------------------------------------------------|-------------------------------------------|
| <b>Client:</b> PARSONS Main of New York, Inc.        | <b>SDG No.:</b> P5246                     |
| <b>Project:</b> Con Edison Non-MGP - Atlantic Avenue | <b>Sample ID:</b> P5246-01                |
| <b>Client ID:</b> MW7-20241209MSD                    | <b>Percent Solids for Spike Sample:</b> 0 |

| Analyte             | Units | Acceptance Limit | Sample Result | Conc. Qualifier | Duplicate Result | Conc. Qualifier | Dilution Factor | RPD/AD | Qual | Analysis Date |
|---------------------|-------|------------------|---------------|-----------------|------------------|-----------------|-----------------|--------|------|---------------|
| Bromide             | mg/L  | +/-20            | 10.1          |                 | 10.1             |                 | 1               | 0      |      | 12/11/2024    |
| Fluoride            | mg/L  | +/-20            | 2.00          |                 | 2.00             |                 | 1               | 0      |      | 12/11/2024    |
| Nitrite             | mg/L  | +/-20            | 2.90          |                 | 2.90             |                 | 1               | 0      |      | 12/11/2024    |
| Chloride            | mg/L  | +/-20            | 390           | OR              | 390              | OR              | 1               | 0      |      | 12/11/2024    |
| Nitrate             | mg/L  | +/-20            | 11.7          | OR              | 11.7             | OR              | 1               | 0      |      | 12/11/2024    |
| Sulfate             | mg/L  | +/-20            | 80.6          | OR              | 80.7             | OR              | 1               | 0      |      | 12/11/2024    |
| Orthophosphate as P | mg/L  | +/-20            | 4.50          |                 | 4.60             |                 | 1               | 2      |      | 12/11/2024    |