

Laboratory Control Sample Summary

Client:	PARSONS Main of New York, Inc.	SDG No.:	P5246
Project:	Con Edison Non-MGP - Atlantic Avenue	Run No.:	LB133881

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133881BSW							
Bromide	mg/L	10	10.1		101	1	90-110	12/11/2024
Chloride	mg/L	3	3.00		100	1	90-110	12/11/2024
Fluoride	mg/L	2	2.00		100	1	90-110	12/11/2024
Nitrite	mg/L	3	3.00		100	1	90-110	12/11/2024
Nitrate	mg/L	2.5	2.50		100	1	90-110	12/11/2024
Sulfate	mg/L	15	14.3		95	1	90-110	12/11/2024
Orthophosphate as P	mg/L	5	4.80		96	1	90-110	12/11/2024

Laboratory Control Sample Summary

Client: PARSONS Main of New York, Inc.

SDG No.: P5246

Project: Con Edison Non-MGP - Atlantic Avenue

Run No.: LB133890

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133890BS							
TDS	mg/L	100	94.0		94	1	90-110	12/10/2024