

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: P5246 SAS No.: P5246 SDG No.: P5246

Instrument ID: MSVOA_X Calibration Date/Time: 12/10/2024 10:43

Lab File ID: VX044197.D Init. Calib. Date(s): 11/21/2024 11/21/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:47 12:03

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.633	0.527		-16.75	20
Chloromethane	0.667	0.569	0.1	-14.69	20
Vinyl Chloride	0.709	0.628		-11.43	20
Bromomethane	0.436	0.455		4.36	20
Chloroethane	0.432	0.424		-1.85	20
Trichlorofluoromethane	1.268	1.418		11.83	20
1,1,2-Trichlorotrifluoroethane	0.605	0.606		0.17	20
1,1-Dichloroethene	0.576	0.529		-8.16	20
Acetone	0.394	0.458		16.24	20
Carbon Disulfide	1.188	0.962		-19.02	20
Methyl tert-butyl Ether	2.092	2.233		6.74	20
Methyl Acetate	0.917	1.022		11.45	20
Methylene Chloride	0.668	0.637		-4.64	20
trans-1,2-Dichloroethene	0.595	0.573		-3.7	20
1,1-Dichloroethane	1.161	1.178	0.1	1.46	20
Cyclohexane	0.956	0.924		-3.35	20
2-Butanone	0.508	0.601		18.31	20
Carbon Tetrachloride	0.500	0.529		5.8	20
cis-1,2-Dichloroethene	0.735	0.749		1.9	20
Bromochloromethane	0.511	0.489		-4.3	20
Chloroform	1.249	1.296		3.76	20
1,1,1-Trichloroethane	1.077	1.143		6.13	20
Methylcyclohexane	0.578	0.582		0.69	20
Benzene	1.387	1.399		0.87	20
1,2-Dichloroethane	0.557	0.598		7.36	20
Trichloroethene	0.393	0.358		-8.91	20
1,2-Dichloropropane	0.340	0.365		7.35	20
Bromodichloromethane	0.482	0.545		13.07	20
4-Methyl-2-Pentanone	0.537	0.657		22.35	20
Toluene	0.830	0.892		7.47	20
t-1,3-Dichloropropene	0.491	0.566		15.27	20
cis-1,3-Dichloropropene	0.539	0.591		9.65	20
1,1,2-Trichloroethane	0.343	0.382		11.37	20
2-Hexanone	0.403	0.515		27.79	20
Dibromochloromethane	0.344	0.401		16.57	20
1,2-Dibromoethane	0.346	0.404		16.76	20
Tetrachloroethene	0.330	0.336		1.82	20
Chlorobenzene	1.098	1.144	0.3	4.19	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: P5246 SAS No.: P5246 SDG No.: P5246

Instrument ID: MSVOA_X Calibration Date/Time: 12/10/2024 10:43

Lab File ID: VX044197.D Init. Calib. Date(s): 11/21/2024 11/21/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:47 12:03

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.861	2.022		8.65	20
m/p-Xylenes	0.694	0.751		8.21	20
o-Xylene	0.678	0.748		10.32	20
Styrene	1.121	1.260		12.4	20
Bromoform	0.240	0.284	0.1	18.33	20
Isopropylbenzene	3.843	4.048		5.33	20
1,1,2,2-Tetrachloroethane	1.316	1.422	0.3	8.06	20
1,3-Dichlorobenzene	1.670	1.754		5.03	20
1,4-Dichlorobenzene	1.702	1.759		3.35	20
1,2-Dichlorobenzene	1.685	1.803		7	20
1,2-Dibromo-3-Chloropropane	0.264	0.303		14.77	20
1,2,4-Trichlorobenzene	0.994	1.061		6.74	20
1,2,3-Trichlorobenzene	1.006	1.115		10.84	20
1,2-Dichloroethane-d4	0.858	0.878		2.33	20
Dibromofluoromethane	0.357	0.374		4.76	20
Toluene-d8	1.232	1.265		2.68	20
4-Bromofluorobenzene	0.419	0.460		9.78	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.