

Surrogate Summary

SDG No.: P5246

Client: PARSONS Main of New York, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5246-01	MW7-20241209	1,2-Dichloroethane-d4	50	52.9	106	74	125
		Dibromofluoromethane	50	47.2	94	75	124
		Toluene-d8	50	49.8	100	86	113
		4-Bromofluorobenzene	50	49.8	100	77	121
P5246-02	TB-20241209	1,2-Dichloroethane-d4	50	54.4	109	74	125
		Dibromofluoromethane	50	46.1	92	75	124
		Toluene-d8	50	49.6	99	86	113
		4-Bromofluorobenzene	50	50.3	101	77	121
VX1210WBL01	VX1210WBL01	1,2-Dichloroethane-d4	50	49.7	99	74	125
		Dibromofluoromethane	50	46.4	93	75	124
		Toluene-d8	50	49.1	98	86	113
		4-Bromofluorobenzene	50	47.5	95	77	121
VX1210WBS01	VX1210WBS01	1,2-Dichloroethane-d4	50	52.6	105	74	125
		Dibromofluoromethane	50	50.9	102	75	124
		Toluene-d8	50	49.6	99	86	113
		4-Bromofluorobenzene	50	51.7	103	77	121
VX1210WBSD0	VX1210WBSD01	1,2-Dichloroethane-d4	50	51.3	103	74	125
		Dibromofluoromethane	50	51.1	102	75	124
		Toluene-d8	50	49.7	99	86	113
		4-Bromofluorobenzene	50	53.8	108	77	121