

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: P5246 SAS No.: P5246 SDG NO.: P5246
Lab File ID: VX044197.D Date Analyzed: 12/10/2024
Instrument ID: MSVOA_X Time Analyzed: 10:43
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	121082	5.54	212067	6.76	187191	10.05
UPPER LIMIT	242164	6.038	424134	7.257	374382	10.549
LOWER LIMIT	60541	5.038	106034	6.257	93595.5	9.549
EPA SAMPLE NO.						
MW7-20241209	107458	5.55	207402	6.76	185141	10.06
TB-20241209	97758	5.55	191902	6.76	171864	10.06
VX1210WBL01	110263	5.54	207593	6.76	178508	10.06
VX1210WBS01	143461	5.54	255477	6.76	217630	10.06
VX1210WBSD01	115787	5.54	202181	6.76	181459	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.