

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51821 MA No.: SDG No.: MJNKG0
Instrument ID: P7 Start Date: 12/21/2024
Analytical Method: ICP-MS Run Batch: LB134060
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	0	0.0	0	1	1.2	20	50	52	4
Copper	0	0.0	0	2	2.1	5	500	520	4
Lead	0	0.0	0	1	0.97	-3	250	250	0
Zinc	0	0.0	0	5	5.5	10	500	490	-2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51821 MA No.: SDG No.: MJNKG0
Instrument ID: P7 Start Date: 12/21/2024
Analytical Method: ICP-MS Run Batch: LB134060
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	125	120	-4	250	250	0	500	500	0
Copper	1250	1200	-4	2500	2600	4	5000	5000	0
Lead	625	580	-7	1250	1300	4	2500	2500	0
Zinc	1250	1100	-12	2500	2600	4	5000	5100	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51821 MA No.: _____ SDG No.: MJNKG0
Instrument ID: P7 Start Date: 12/21/2024
Analytical Method: ICP-MS Run Batch: LB134060
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	1000	1000	0						
Copper	10000	10000	0						
Lead	5000	5000	0						
Zinc	10000	9900	-1						