

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51821 MA No. : SDG No.: MJNKG0
 Initial Calibration Verification Source : MP83629
 Continuing Calibration Verification Source : MP83632
 Run Batch: LB134060 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV006				ID: CCV037				ID: CCV038		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	200	220	108	1	500	490	98	0	480	96	1
Copper	100	98	98	0	5000	4700	95	1	4800	95	1
Lead	200	200	100	1	2500	2600	102	1	2500	101	1
Zinc	200	200	101	1	5000	4700	95	2	4600	93	1