

FORM 4 - IN
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51821 MA No. : SDG No.: MJNKG0
 Analytical Method: ICP-MS ICSA Source: EPA-0803
 Instrument ID: P7 ICSB Source: EPA-0803
 Run Batch: LB134060
 Concentration Units: ug/L

Analyte	True		Found			
	ICSA	ICSAB	ICSA	%R	ICSAB	%R
Aluminum	100000	100000	92000	92	91000	91
Antimony	1.5	22	1.1	74	21	96
Arsenic	0.1	19	0.31	310	21	111
Barium	1.2	22	1.4	118	22	98
Beryllium	0	19	0.3		21	113
Cadmium	0.7	20	0.33	47	17	87
Calcium	100000	100000	100000	102	100000	101
Chromium	21	40	21	98	41	102
Cobalt	1	20	1.3	125	21	107
Copper	8	25	7.1	89	26	105
Iron	100000	100000	100000	104	100000	103
Lead	4	25	4.8	119	24	95
Magnesium	100000	100000	95000	95	94000	93
Manganese	7	27	7.9	113	29	106
Molybdenum	2000	2000	2000	99	2000	98
Nickel	6	24	5.4	90	25	106
Potassium	100000	100000	98000	98	97000	97
Selenium	0.3	19	0.4	133	21	111
Silver	0	18	0.05		19	105
Sodium	100000	100000	99000	99	99000	98
Thallium	0	21	0.14		20	94
Vanadium	0.5	19	0.2	40	21	108
Zinc	11	29	11	99	31	108