

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51918 MA No. : SDG No.: MJNBZ1
Initial Calibration Verification Source : MP83793
Continuing Calibration Verification Source : MP83795
Run Batch: LB134105 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV033				ID: CCV074				ID: CCV075		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.7	93		5.0	4.6	91		4.8	97	