

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134096

Review By	Sarabjit Jaswal	Review On	12/30/2024 2:19:55 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 2:20:35 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	
LCS Standard	MP83669
Chk Standard	MP83635,MP83634

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE001	TUNE	12/26/24 13:05		Jaswal	OK
2	S00	S0	CAL1	12/26/24 13:42		Jaswal	OK
3	S02	S02	CAL3	12/26/24 13:48		Jaswal	OK
4	S03	S03	CAL4	12/26/24 13:52		Jaswal	OK
5	S04	S04	CAL5	12/26/24 13:55		Jaswal	OK
6	S05	S05	CAL6	12/26/24 13:58		Jaswal	OK
7	S06	S06	CAL7	12/26/24 14:00		Jaswal	OK
8	S07	S07	CAL8	12/26/24 14:03		Jaswal	OK
9	S08	S08	CAL9	12/26/24 14:06		Jaswal	OK
10	ICV001	ICV001	ICV	12/26/24 14:32	Inst. paused for calibration review	Jaswal	OK
11	ICB001	ICB001	ICB	12/26/24 14:42	Inst. paused for review	Jaswal	OK
12	ICSA001	ICSA001	ICSA	12/26/24 14:45		Jaswal	OK
13	ICSAB001	ICSAB001	ICSAB	12/26/24 14:48		Jaswal	OK
14	CCV001	CCV001	CCV	12/26/24 14:54	Inst. paused for review	Jaswal	OK
15	CCB001	CCB001	CCB	12/26/24 15:00	Inst. paused for review	Jaswal	OK
16	PB165755BL	PBS755	MB	12/26/24 15:03		Jaswal	OK
17	PB165755BS	LCS755	LCS	12/26/24 15:07		Jaswal	OK
18	PB165818BL	PBS818	MB	12/26/24 15:10		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134096

Review By	Sarabjit Jaswal	Review On	12/30/2024 2:19:55 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 2:20:35 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard	MP83669		
Chk Standard	MP83635,MP83634		

19	PB165818BS	LCS818	LCS	12/26/24 15:13		Jaswal	OK
20	PB165819BL	PBS819	MB	12/26/24 15:17		Jaswal	OK
21	PB165819BS	LCS819	LCS	12/26/24 15:20		Jaswal	OK
22	PB165760BL	PBS760	MB	12/26/24 15:23		Jaswal	OK
23	PB165760BS	LCS760	LCS	12/26/24 15:27		Jaswal	OK
24	P5255-01	MJNL27	SAM	12/26/24 15:30		Jaswal	OK
25	P5255-02	MJNL72	SAM	12/26/24 15:33		Jaswal	OK
26	P5255-03	MJNL72D	DUP	12/26/24 15:36		Jaswal	OK
27	P5255-02L	MJNL72L	SD	12/26/24 15:39		Jaswal	OK
28	P5255-04	MJNL72S	MS	12/26/24 15:42	MS Fail For As(25.32ppb)	Jaswal	OK
29	P5255-05	MJNL93	SAM	12/26/24 15:46		Jaswal	OK
30	P5255-06	MJNL94	SAM	12/26/24 15:49		Jaswal	OK
31	P5256-01	MJNL92	SAM	12/26/24 15:52		Jaswal	OK
32	P5256-02	MJNL92D	DUP	12/26/24 15:55		Jaswal	OK
33	P5256-01L	MJNL92L	SD	12/26/24 15:58		Jaswal	OK
34	P5256-03	MJNL92S	MS	12/26/24 16:01		Jaswal	OK
35	PB165821BL	PBS821	MB	12/26/24 16:04	Not Use	Jaswal	Not Ok
36	PB165821BS	LCS821	LCS	12/26/24 16:07	Not Use	Jaswal	Not Ok
37	P5250-06	MJNBZ2	SAM	12/26/24 16:11	Not Use	Jaswal	Not Ok
38	P5250-07	MJNBZ2D	DUP	12/26/24 16:14	Not Use	Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134096

Review By	Sarabjit Jaswal	Review On	12/30/2024 2:19:55 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 2:20:35 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	
LCS Standard	MP83669
Chk Standard	MP83635,MP83634

39	P5250-06L	MJNBZ2L	SD	12/26/24 16:17	Not Use	Jaswal	Not Ok
40	P5250-08	MJNBZ2S	MS	12/26/24 16:20	Not Use	Jaswal	Not Ok
41	CCV002	CCV002	CCV	12/26/24 16:23		Jaswal	OK
42	CCB002	CCB002	CCB	12/26/24 16:30	Inst. paused for review	Jaswal	OK
43	P5271-06	MJNKC7	SAM	12/26/24 16:33	As High	Jaswal	Dilution
44	P5271-07	MJNKC7D	DUP	12/26/24 16:36	As High	Jaswal	Dilution
45	P5271-06L	MJNKC7L	SD	12/26/24 16:39	As High	Jaswal	Dilution
46	P5271-08	MJNKC7S	MS	12/26/24 16:42	As High	Jaswal	Dilution
47	P5266-18RE	ME2903	SAM	12/26/24 16:45	Not Use	Jaswal	Not Ok
48	P5266-19RE	ME28Y9D	DUP	12/26/24 16:48	Not Use	Jaswal	Not Ok
49	P5266-04LRE	ME28Y9L	SD	12/26/24 16:51	Not Use	Jaswal	Not Ok
50	P5266-20RE	ME28Y9S	MS	12/26/24 16:54	Not Use	Jaswal	Not Ok
51	P5259-04	ME28X0	SAM	12/26/24 16:57		Jaswal	OK
52	P5259-05	ME28X0D	DUP	12/26/24 17:00		Jaswal	OK
53	P5259-04L	ME28X0L	SD	12/26/24 17:04		Jaswal	OK
54	P5259-06	ME28X0S	MS	12/26/24 17:07	MS Fail For AS(17.12ppb),(Sb,Cd,Se,Tl -Below RL)	Jaswal	OK
55	P5266-01	ME28Y6	SAM	12/26/24 17:10		Jaswal	OK
56	P5266-02	ME28Y7	SAM	12/26/24 17:13		Jaswal	OK
57	P5266-03	ME28Y8	SAM	12/26/24 17:16		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134096

Review By	Sarabjit Jaswal	Review On	12/30/2024 2:19:55 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 2:20:35 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard	MP83669		
Chk Standard	MP83635,MP83634		

58	P5266-04	ME28Y9	SAM	12/26/24 17:19	Not USe	Jaswal	Not Ok
59	P5266-05	ME28Z0	SAM	12/26/24 17:22		Jaswal	OK
60	P5266-06	ME28Z1	SAM	12/26/24 17:26		Jaswal	OK
61	P5266-07	ME28Z2	SAM	12/26/24 17:29		Jaswal	OK
62	P5266-08	ME28Z3	SAM	12/26/24 17:32		Jaswal	OK
63	P5266-09	ME28Z4	SAM	12/26/24 17:35		Jaswal	OK
64	P5266-10	ME28Z5	SAM	12/26/24 17:38		Jaswal	OK
65	P5266-11	ME28Z6	SAM	12/26/24 17:41		Jaswal	OK
66	P5266-12	ME28Z7	SAM	12/26/24 17:44		Jaswal	OK
67	P5266-13	ME28Z8	SAM	12/26/24 17:47		Jaswal	OK
68	CCV003	CCV003	CCV	12/26/24 17:51		Jaswal	OK
69	CCB003	CCB003	CCB	12/26/24 17:53		Jaswal	OK
70	P5266-14	ME28Z9	SAM	12/26/24 17:58		Jaswal	OK
71	P5266-15	ME2900	SAM	12/26/24 18:01		Jaswal	OK
72	P5266-16	ME2901	SAM	12/26/24 18:05		Jaswal	OK
73	P5259-01	ME28W7	SAM	12/26/24 18:11		Jaswal	OK
74	P5259-02	ME28W8	SAM	12/26/24 18:14		Jaswal	OK
75	P5259-03	ME28W9	SAM	12/26/24 18:17		Jaswal	OK
76	P5259-07	ME28X1	SAM	12/26/24 18:20		Jaswal	OK
77	P5259-08	ME28X2	SAM	12/26/24 18:23		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134096

Review By	Sarabjit Jaswal	Review On	12/30/2024 2:19:55 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 2:20:35 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard	MP83669		
Chk Standard	MP83635,MP83634		

78	P5259-09	ME28X3	SAM	12/26/24 18:27		Jaswal	OK
79	P5259-10	ME28X4	SAM	12/26/24 18:30		Jaswal	OK
80	P5259-11	ME28X5	SAM	12/26/24 18:33		Jaswal	OK
81	P5259-12	ME28X6	SAM	12/26/24 18:36		Jaswal	OK
82	P5259-13	ME28X7	SAM	12/26/24 18:39		Jaswal	OK
83	P5259-14	ME28X8	SAM	12/26/24 18:42		Jaswal	OK
84	P5259-15	ME28X9	SAM	12/26/24 18:45		Jaswal	OK
85	P5259-16	ME28Y0	SAM	12/26/24 18:49		Jaswal	OK
86	P5259-17	ME28Y1	SAM	12/26/24 18:52		Jaswal	OK
87	P5259-18	ME28Y2	SAM	12/26/24 18:55		Jaswal	OK
88	P5259-19	ME28Y3	SAM	12/26/24 18:58		Jaswal	OK
89	P5259-20	ME28Y4	SAM	12/26/24 19:01		Jaswal	OK
90	P5259-21	ME28Y5	SAM	12/26/24 19:04		Jaswal	OK
91	P5271-01	MJNKB7	SAM	12/26/24 19:07		Jaswal	OK
92	P5271-02	MJNKB8	SAM	12/26/24 19:11		Jaswal	OK
93	P5271-03	MJNKC0	SAM	12/26/24 19:14		Jaswal	OK
94	P5271-04	MJNKC5	SAM	12/26/24 19:17	As High	Jaswal	Dilution
95	P5271-05	MJNKC6	SAM	12/26/24 19:20		Jaswal	OK
96	CCV004	CCV004	CCV	12/26/24 19:26		Jaswal	OK
97	CCB004	CCB004	CCB	12/26/24 19:28		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134096

Review By	Sarabjit Jaswal	Review On	12/30/2024 2:19:55 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 2:20:35 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard	MP83669		
Chk Standard	MP83635,MP83634		

98	P5271-09	MJNKC8	SAM	12/26/24 19:32	As High	Jaswal	Dilution
99	P5271-10	MJNKC9	SAM	12/26/24 19:35		Jaswal	OK
100	P5271-11	MJNKH2	SAM	12/26/24 19:38		Jaswal	OK
101	P5271-12	MJNKH3	SAM	12/26/24 19:41		Jaswal	OK
102	P5271-13	MJNKH4	SAM	12/26/24 19:44		Jaswal	OK
103	P5271-14	MJNKH5	SAM	12/26/24 19:47		Jaswal	OK
104	P5271-15	MJNKH6	SAM	12/26/24 19:50		Jaswal	OK
105	P5271-16	MJNKH7	SAM	12/26/24 19:53		Jaswal	OK
106	P5271-17	MJNKH8	SAM	12/26/24 19:57		Jaswal	OK
107	P5271-18	MJNKKJ2	SAM	12/26/24 20:00		Jaswal	OK
108	P5271-19	MJNKKJ3	SAM	12/26/24 20:03		Jaswal	OK
109	P5271-20	MJNKKJ4	SAM	12/26/24 20:06		Jaswal	OK
110	P5271-21	MJNKKJ5	SAM	12/26/24 20:09		Jaswal	OK
111	P5271-06DL	MJNKC7	SAM	12/26/24 20:12	2X For As	Jaswal	Confirms
112	P5271-07DL	MJNKC7D	DUP	12/26/24 20:15	2X For As	Jaswal	Confirms
113	P5271-06LDL	MJNKC7L	SD	12/26/24 20:18	10X For As	Jaswal	Confirms
114	P5271-08DL	MJNKC7S	MS	12/26/24 20:21	2X For As	Jaswal	Confirms
115	P5266-18	ME2903	SAM	12/26/24 20:24	Not Use	Jaswal	Not Ok
116	P5266-19	ME28Y9D	DUP	12/26/24 20:27	Not Use	Jaswal	Not Ok
117	P5266-04L	ME28Y9L	SD	12/26/24 20:30	Not Use	Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134096

Review By	Sarabjit Jaswal	Review On	12/30/2024 2:19:55 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 2:20:35 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard	MP83669		
Chk Standard	MP83635,MP83634		

118	P5266-20	ME28Y9S	MS	12/26/24 20:34	Not Use	Jaswal	Not Ok
119	P5266-18DL	ME2903	SAM	12/26/24 20:37	Not Use	Jaswal	Not Ok
120	P5266-19DL	ME28Y9D	DUP	12/26/24 20:40	Not Use	Jaswal	Not Ok
121	P5266-04LDL	ME28Y9L	SD	12/26/24 20:43	Not Use	Jaswal	Not Ok
122	P5266-20DL	ME28Y9S	MS	12/26/24 20:46	Not Use	Jaswal	Not Ok
123	P5271-04DL	MJNKC5	SAM	12/26/24 20:49	5X For As	Jaswal	Confirms
124	CCV005	CCV005	CCV	12/26/24 20:52		Jaswal	OK
125	CCB005	CCB005	CCB	12/26/24 20:55		Jaswal	OK
126	P5255-02A	MJNL72A	PS	12/26/24 20:58	Not Use	Jaswal	Not Ok
127	P5256-01A	MJNL92A	PS	12/26/24 21:01	Not Use	Jaswal	Not Ok
128	P5259-04A	ME28X0A	PS	12/26/24 21:04	Not Use	Jaswal	Not Ok
129	P5271-06A	MJNKC7A	PS	12/26/24 21:07	Not Use	Jaswal	Not Ok
130	P5271-09DL	MJNKC8	SAM	12/26/24 21:10	2X for As	Jaswal	Confirms
131	CCV006	CCV006	CCV	12/26/24 21:13		Jaswal	OK
132	CCB006	CCB006	CCB	12/26/24 21:16		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134112

Review By	Sarabjit Jaswal	Review On	12/30/2024 3:27:20 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 3:28:46 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard	MP83764		
LCS Standard	MP83670		
Chk Standard	MP83635,MP83634		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE002	TUNE	12/27/24 14:06		Jaswal	OK
2	S00	S0	CAL1	12/27/24 14:36		Jaswal	OK
3	S01	S01	CAL2	12/27/24 14:46		Jaswal	OK
4	S02	S02	CAL3	12/27/24 14:49		Jaswal	OK
5	S03	S03	CAL4	12/27/24 14:52		Jaswal	OK
6	S04	S04	CAL5	12/27/24 14:55		Jaswal	OK
7	S05	S05	CAL6	12/27/24 14:58		Jaswal	OK
8	S06	S06	CAL7	12/27/24 15:01		Jaswal	OK
9	S07	S07	CAL8	12/27/24 15:04		Jaswal	OK
10	S08	S08	CAL9	12/27/24 15:07		Jaswal	OK
11	ICV002	ICV002	ICV	12/27/24 15:30	Inst. paused for calibration review	Jaswal	OK
12	ICB002	ICB002	ICB	12/27/24 15:38	Inst. paused for review	Jaswal	OK
13	ICSA002	ICSA002	ICSA	12/27/24 15:42		Jaswal	OK
14	ICSAB002	ICSAB002	ICSAB	12/27/24 15:45		Jaswal	OK
15	CCV011	CCV011	CCV	12/27/24 15:48		Jaswal	OK
16	CCB011	CCB011	CCB	12/27/24 15:51		Jaswal	OK
17	P5306-02DL	OU4-VSL-07-121224	SAM	12/27/24 15:57	Not Use	Jaswal	Not Ok
18	P5306-02DUPDL	OU4-VSL-07-121224	DUP	12/27/24 16:01	Not Use	Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134112

Review By	Sarabjit Jaswal	Review On	12/30/2024 3:27:20 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 3:28:46 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	MP83764
LCS Standard	MP83670
Chk Standard	MP83635,MP83634

19	P5306-02LDL	OU4-VSL-07-121224	SD	12/27/24 16:04	Not Use	Jaswal	Not Ok
20	P5306-02MSDL	OU4-VSL-07-121224	MSD	12/27/24 16:07	Not Use	Jaswal	Not Ok
21	P5306-02MSDDL	OU4-VSL-07-121224	MSD	12/27/24 16:10	Not Use	Jaswal	Not Ok
22	P5306-02ADL	OU4-VSL-07-121224	PS	12/27/24 16:13	Not Use	Jaswal	Not Ok
23	P5306-08DL	OU4-VSL-10-121224	SAM	12/27/24 16:16	Not Use	Jaswal	Not Ok
24	P5306-10DL	OU4-VSL-11-121224	SAM	12/27/24 16:19	Not Use	Jaswal	Not Ok
25	CCV012	CCV012	CCV	12/27/24 16:23		Jaswal	OK
26	CCB012	CCB012	CCB	12/27/24 16:25		Jaswal	OK
27	PB165821BL	PBS821	MB	12/27/24 16:29		Jaswal	OK
28	PB165821BS	LCS821	LCS	12/27/24 16:32		Jaswal	OK
29	P5255-02A	MJNL72A	PS	12/27/24 16:35	PS For As-0.05-MP83764 - 10ml Sample	Jaswal	OK
30	P5256-01ARE	MJNL92A	PS	12/27/24 16:41	Not Use	Jaswal	Not Ok
31	P5259-04A	ME28X0A	PS	12/27/24 16:44	PS For Sb,As,Cd,Se,Tl	Jaswal	OK
32	P5271-06A	MJNKC7A	PS	12/27/24 16:47	Not Required	Jaswal	Not Ok
33	P5266-04	ME28Y9	SAM	12/27/24 16:50		Jaswal	OK
34	P5266-19	ME28Y9D	DUP	12/27/24 16:53		Jaswal	OK
35	P5266-04LDL	ME28Y9L	SD	12/27/24 16:57		Jaswal	OK
36	P5266-20	ME28Y9S	MS	12/27/24 17:00	MS Fail For Sb,As,Cd,Se,Tl,Zn	Jaswal	OK
37	P5266-17	ME2902	SAM	12/27/24 17:03		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134112

Review By	Sarabjit Jaswal	Review On	12/30/2024 3:27:20 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 3:28:46 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	MP83764
LCS Standard	MP83670
Chk Standard	MP83635,MP83634

38	P5266-18	ME2903	SAM	12/27/24 17:06	Cr High	Jaswal	Dilution
39	P5256-01ARE2	MJNL92A	PS	12/27/24 17:09	Not USe	Jaswal	Not Ok
40	P5250-01	MJNBZ1	SAM	12/27/24 17:12		Jaswal	OK
41	P5250-02	MJNBZ3	SAM	12/27/24 17:16		Jaswal	OK
42	P5256-01A	MJNL92A	PS	12/27/24 17:20		Jaswal	Not Ok
43	P5250-03	MJNBZ4	SAM	12/27/24 17:23		Jaswal	OK
44	P5250-04	MJNBZ7	SAM	12/27/24 17:26		Jaswal	OK
45	P5250-05	MJNBZ8	SAM	12/27/24 17:29		Jaswal	OK
46	P5250-06	MJNBZ2	SAM	12/27/24 17:33		Jaswal	OK
47	P5250-07	MJNBZ2D	DUP	12/27/24 17:36		Jaswal	OK
48	P5250-06L	MJNBZ2L	SD	12/27/24 17:39		Jaswal	OK
49	P5250-08	MJNBZ2S	MS	12/27/24 17:42	MS Fail For Sb(Below RL),As(9.80ppb),Cd(Below RL),Mo(Below RL),Se(Below RL),Ti(Below RL)	Jaswal	OK
50	PB165880BL	PBS880	MB	12/27/24 17:45		Jaswal	OK
51	PB165880BS	LCS880	LCS	12/27/24 17:49		Jaswal	OK
52	PB165859BL	PBW859	MB	12/27/24 17:52		Jaswal	OK
53	PB165859BS	LCS859	LCS	12/27/24 17:55		Jaswal	OK
54	CCV013	CCV013	CCV	12/27/24 18:01		Jaswal	OK
55	CCB013	CCB013	CCB	12/27/24 18:11	Inst. paused for review	Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134112

Review By	Sarabjit Jaswal	Review On	12/30/2024 3:27:20 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 3:28:46 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	MP83764
LCS Standard	MP83670
Chk Standard	MP83635,MP83634

56	PB165867BL	PBS867	MB	12/27/24 18:14		Jaswal	OK
57	PB165867BS	LCS867	LCS	12/27/24 18:18		Jaswal	OK
58	PB165866BL	PBS866	MB	12/27/24 18:21		Jaswal	OK
59	PB165866BS	LCS866	LCS	12/27/24 18:24		Jaswal	OK
60	P5347-05	MJNBZ5	SAM	12/27/24 18:28		Jaswal	OK
61	P5347-06	MJNBZ5D	DUP	12/27/24 18:31		Jaswal	OK
62	P5347-05L	MJNBZ5L	SD	12/27/24 18:34		Jaswal	OK
63	P5347-07	MJNBZ5S	MS	12/27/24 18:37	MS Fail For Sb,As,Cd,Se,Tl	Jaswal	OK
64	P5364-06	MJNKB1	SAM	12/27/24 18:40	As,Zn High	Jaswal	Dilution
65	P5364-07	MJNKB1D	DUP	12/27/24 18:43	As,Zn High	Jaswal	Dilution
66	P5364-06L	MJNKB1L	SD	12/27/24 18:46	As,Zn High	Jaswal	Dilution
67	P5364-08	MJNKB1S	MS	12/27/24 18:49	As,Zn High	Jaswal	Dilution
68	P5352-04	ME2926	SAM	12/27/24 18:52		Jaswal	OK
69	P5352-05	ME2926D	DUP	12/27/24 18:56		Jaswal	OK
70	P5352-04L	ME2926L	SD	12/27/24 18:59		Jaswal	OK
71	P5352-06	ME2926S	MS	12/27/24 19:02	MS Fail For Sb,As,Cd,Se,Tl	Jaswal	OK
72	P5363-06	MJNKA6	SAM	12/27/24 19:05		Jaswal	OK
73	P5363-07	MJNKA6D	DUP	12/27/24 19:08		Jaswal	OK
74	P5363-06L	MJNKA6L	SD	12/27/24 19:11		Jaswal	OK
75	P5363-06RE	MJNKA6	SAM	12/27/24 19:15	Not Use	Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134112

Review By	Sarabjit Jaswal	Review On	12/30/2024 3:27:20 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 3:28:46 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard	MP83764		
LCS Standard	MP83670		
Chk Standard	MP83635,MP83634		

76	P5347-01	MJNBZ9	SAM	12/27/24 19:18		Jaswal	OK
77	CCV014	CCV014	CCV	12/27/24 19:21		Jaswal	OK
78	CCB014	CCB014	CCB	12/27/24 19:24		Jaswal	OK
79	P5347-02	MJNBZ9D	DUP	12/27/24 19:27		Jaswal	OK
80	P5347-01L	MJNBZ9L	SD	12/27/24 19:31		Jaswal	OK
81	P5347-03	MJNBZ9S	MS	12/27/24 19:34		Jaswal	OK
82	P5354-02	ME2928	SAM	12/27/24 19:37		Jaswal	OK
83	P5354-03	ME2928D	DUP	12/27/24 19:40		Jaswal	OK
84	P5354-02L	ME2928L	SD	12/27/24 19:44		Jaswal	OK
85	P5354-04	ME2928S	MS	12/27/24 19:47		Jaswal	OK
86	P5254-12	MJNLG1	SAM	12/27/24 19:50		Jaswal	OK
87	P5255-07	MJNLC3	SAM	12/27/24 19:54		Jaswal	OK
88	P5347-04	MJNC00	SAM	12/27/24 19:57		Jaswal	OK
89	P5354-01	ME2927	SAM	12/27/24 20:00		Jaswal	OK
90	P5363-14	MJNLE3	SAM	12/27/24 20:04		Jaswal	OK
91	P5363-15	MJNLF9	SAM	12/27/24 20:07		Jaswal	OK
92	P5363-16	MJNLG0	SAM	12/27/24 20:10		Jaswal	OK
93	P5364-20	MJNLC4	SAM	12/27/24 20:13		Jaswal	OK
94	P5364-01	MJNKA1	SAM	12/27/24 20:17		Jaswal	OK
95	P5364-02	MJNKA7	SAM	12/27/24 20:20		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134112

Review By	Sarabjit Jaswal	Review On	12/30/2024 3:27:20 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 3:28:46 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	MP83764
LCS Standard	MP83670
Chk Standard	MP83635,MP83634

96	P5364-03	MJNKA8	SAM	12/27/24 20:23		Jaswal	OK
97	P5364-04	MJNKA9	SAM	12/27/24 20:26		Jaswal	OK
98	P5364-05	MJNKB0	SAM	12/27/24 20:29	As High	Jaswal	Dilution
99	P5364-09	MJNKB2	SAM	12/27/24 20:32	As,Zn High	Jaswal	Dilution
100	P5364-10	MJNKB3	SAM	12/27/24 20:36		Jaswal	OK
101	P5364-11	MJNKB4	SAM	12/27/24 20:39	As,Zn High	Jaswal	Dilution
102	P5364-12	MJNKB5	SAM	12/27/24 20:42		Jaswal	OK
103	P5364-13	MJNKB9	SAM	12/27/24 20:45		Jaswal	OK
104	P5364-14	MJNKC1	SAM	12/27/24 20:49		Jaswal	OK
105	P5364-15	MJNKC2	SAM	12/27/24 20:52		Jaswal	OK
106	CCV015	CCV015	CCV	12/27/24 20:55		Jaswal	OK
107	CCB015	CCB015	CCB	12/27/24 21:06	Inst. paused for review	Jaswal	OK
108	P5364-16	MJNKC3	SAM	12/27/24 21:09		Jaswal	OK
109	P5364-17	MJNKG8	SAM	12/27/24 21:12		Jaswal	OK
110	P5364-18	MJNKH9	SAM	12/27/24 21:15		Jaswal	OK
111	P5364-19	MJNKA0	SAM	12/27/24 21:19		Jaswal	OK
112	P5352-01	ME2922	SAM	12/27/24 21:22		Jaswal	OK
113	P5352-02	ME2924	SAM	12/27/24 21:25		Jaswal	OK
114	P5352-03	ME2925	SAM	12/27/24 21:28		Jaswal	OK
115	P5363-01	MJNKA0	SAM	12/27/24 21:32		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134112

Review By	Sarabjit Jaswal	Review On	12/30/2024 3:27:20 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 3:28:46 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	MP83764
LCS Standard	MP83670
Chk Standard	MP83635,MP83634

116	P5363-02	MJNKA2	SAM	12/27/24 21:35	As High	Jaswal	Dilution
117	P5363-08	MJNKA6S	MS	12/27/24 21:38		Jaswal	OK
118	P5363-03	MJNKA3	SAM	12/27/24 21:41	As High	Jaswal	Dilution
119	P5363-04	MJNKA4	SAM	12/27/24 21:44	As High	Jaswal	Dilution
120	P5363-05	MJNKA5	SAM	12/27/24 21:47	As,Pb,Zn High , INT_STD 209Bi(1) Fail	Jaswal	Dilution
121	P5363-09	MJNKB6	SAM	12/27/24 21:50	As,Pb,Zn High	Jaswal	Dilution
122	P5363-10	MJNKC4	SAM	12/27/24 21:53		Jaswal	OK
123	P5363-11	MJNKG9	SAM	12/27/24 21:56		Jaswal	OK
124	P5363-12	MJNKH0	SAM	12/27/24 21:59		Jaswal	OK
125	P5363-13	MJNKK0	SAM	12/27/24 22:02		Jaswal	OK
126	CCV016	CCV016	CCV	12/27/24 22:32		Jaswal	OK
127	CCB016	CCB016	CCB	12/27/24 22:35		Jaswal	OK
128	P5364-06ARE	MJNKB1A	PS	12/27/24 22:41	Not Use	Jaswal	Not Ok
129	P5266-04ARE	ME28Y9A	PS	12/27/24 22:43	Not Use	Jaswal	Not Ok
130	P5352-04ARE	ME2926A	PS	12/27/24 22:47	Not Use	Jaswal	Not Ok
131	P5347-05ARE	MJNBZ5A	PS	12/27/24 22:50	Not Use	Jaswal	Not Ok
132	P5250-06ARE	MJNBZ2A	PS	12/27/24 22:53	Not Use	Jaswal	Not Ok
133	P5364-06ARE2	MJNKB1A	PS	12/27/24 22:56	Not Use	Jaswal	Not Ok
134	CCV017	CCV017	CCV	12/27/24 23:44		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134112

Review By	Sarabjit Jaswal	Review On	12/30/2024 3:27:20 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 3:28:46 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard	MP83764		
LCS Standard	MP83670		
Chk Standard	MP83635,MP83634		

135	CCB017	CCB017	CCB	12/27/24 23:47		Jaswal	OK
136	P5266-02DL	ME28Y7	SAM	12/27/24 23:51	Not Required	Jaswal	Not Ok
137	P5364-06DL	MJNKB1	SAM	12/27/24 23:54	2X For As,Zn , Still As High	Jaswal	Dilution
138	P5364-07DL	MJNKB1D	DUP	12/27/24 23:57	2X For As,Zn , Still As High	Jaswal	Dilution
139	P5364-06LDL	MJNKB1L	SD	12/28/24	2X For As,Zn , Still As High	Jaswal	Confirms
140	P5364-08DL	MJNKB1S	MS	12/28/24 00:03	2X For As,Zn , Still As High	Jaswal	Dilution
141	P5364-06DL2	MJNKB1	SAM	12/28/24 00:06	Not Use	Jaswal	Not Ok
142	P5364-07DL2	MJNKB1D	DUP	12/28/24 00:09	Not Use	Jaswal	Not Ok
143	P5364-06LDL2	MJNKB1L	SD	12/28/24 00:12	Not Use	Jaswal	Not Ok
144	P5364-08DL2	MJNKB1S	MS	12/28/24 00:16	Not Use	Jaswal	Not Ok
145	P5364-06DL3	MJNKB1	SAM	12/28/24 00:19	10X For As	Jaswal	Confirms
146	P5364-07DL3	MJNKB1D	DUP	12/28/24 00:22	10X For As	Jaswal	Confirms
147	P5364-06LDL3	MJNKB1L	SD	12/28/24 00:25	50X For As	Jaswal	Confirms
148	P5364-08DL3	MJNKB1S	MS	12/28/24 00:28	10X For As	Jaswal	Confirms
149	P5364-05DL	MJNKB0	SAM	12/28/24 00:32	2X For As	Jaswal	Confirms
150	P5364-09DL	MJNKB2	SAM	12/28/24 00:35	5X For As,Zn , Still As High	Jaswal	Dilution
151	P5364-09DL2	MJNKB2	SAM	12/28/24 00:38	20X for As	Jaswal	Confirms
152	P5364-11DL	MJNKB4	SAM	12/28/24 00:41	2X For As,Zn , Still As High	Jaswal	Dilution
153	P5364-11DL2	MJNKB4	SAM	12/28/24 00:44	Not Use	Jaswal	Not Ok
154	P5364-11DL3	MJNKB4	SAM	12/28/24 00:47	10X For As	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134112

Review By	Sarabjit Jaswal	Review On	12/30/2024 3:27:20 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 3:28:46 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	MP83764
LCS Standard	MP83670
Chk Standard	MP83635,MP83634

155	P5363-02DL	MJNKA2	SAM	12/28/24 00:50	Not Use	Jaswal	Not Ok
156	P5363-02DL2	MJNKA2	SAM	12/28/24 00:54	5X For As	Jaswal	Confirms
157	P5363-03DL	MJNKA3	SAM	12/28/24 00:57	5X For As	Jaswal	Dilution
158	CCV018	CCV018	CCV	12/28/24 01:00		Jaswal	OK
159	CCB018	CCB018	CCB	12/28/24 01:03		Jaswal	OK
160	P5363-04DL	MJNKA4	SAM	12/28/24 01:06	Not Use	Jaswal	Not Ok
161	P5363-04DL2	MJNKA4	SAM	12/28/24 01:09	5X For As	Jaswal	Confirms
162	P5363-05DL	MJNKA5	SAM	12/28/24 01:12	2X For INT_STD 209Bi(1), 2X For AsmPb,Zn, Still High	Jaswal	Dilution
163	P5363-05DL2	MJNKA5	SAM	12/28/24 01:15	10X For As,Pb,Zn , Still As High	Jaswal	Dilution
164	P5363-05DL3	MJNKA5	SAM	12/28/24 01:18	25X For As	Jaswal	Confirms
165	P5363-09DL	MJNKB6	SAM	12/28/24 01:22	2X For As,Pb,Zn , Still As,Zn High	Jaswal	Dilution
166	P5363-09DL2	MJNKB6	SAM	12/28/24 01:25	5X For As,Zn, Still As High	Jaswal	Dilution
167	P5363-09DL3	MJNKB6	SAM	12/28/24 01:28	Not Use	Jaswal	Not Ok
168	P5363-09DL4	MJNKB6	SAM	12/28/24 01:31	25X For As	Jaswal	Confirms
169	P5364-06A	MJNKB1A	PS	12/28/24 01:34	Not Required	Jaswal	Not Ok
170	P5266-04A	ME28Y9A	PS	12/28/24 01:37	PS For Sb,As,Cd,Se,Tl,Zn	Jaswal	OK
171	P5352-04A	ME2926A	PS	12/28/24 01:40	PS For Sb,As,Cd,Se,Tl	Jaswal	OK
172	P5347-05A	MJNBZ5A	PS	12/28/24 01:43	PS For Sb,As,Cd,Se,Tl	Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QCBatch ID # LB134112

Review By	Sarabjit Jaswal	Review On	12/30/2024 3:27:20 AM
Supervise By	Mohan Bera	Supervise On	12/30/2024 3:28:46 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	MP83764
LCS Standard	MP83670
Chk Standard	MP83635,MP83634

173	P5250-06A	MJNBZ2A	PS	12/28/24 01:47	Not Use	Jaswal	Not Ok
174	CCV019	CCV019	CCV	12/28/24 01:50		Jaswal	OK
175	CCB019	CCB019	CCB	12/28/24 01:53		Jaswal	OK