

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51847 MA No. : SDG No.: ME28W7
Initial Calibration Verification Source : WP111102
Continuing Calibration Verification Source : WP111103
Run Batch: LB133980 Analytical Method: Spectrophotometry
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Cyanide	99.0	97	98		250	240	96		260	102	