

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51847 MA No.: SDG No.: ME28W7
Instrument ID: CV1 Start Date: 12/26/2024
Analytical Method: CVAA Run Batch: LB134089
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.047	0	0.2	0.19	-5	2.5	2.6	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51847 MA No.: _____ SDG No.: ME28W7
Instrument ID: CV1 Start Date: 12/26/2024
Analytical Method: CVAA Run Batch: LB134089
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	4.9	-1	7.5	7.3	-2	10	10	1