

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5260

Project: Monthly 2024

Run No.: LB133930

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133930BS							
BOD5	mg/L	198	187		95	1	84.6-115.4	12/13/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5260

Project: Monthly 2024

Run No.: LB133934

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133934BS							
TPH	mg/L	20.0	17.0		85	1	78-114	12/13/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5260

Project: Monthly 2024

Run No.: LB133934

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133934BSD							
TPH	mg/L	20.0	17.3		86	1	78-114	12/13/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5260

Project: Monthly 2024

Run No.: LB133963

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133963BS							
TSS	mg/L	550	510		93	1	90-110	12/16/2024