

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/10/24
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/11/24
Client Sample ID:	RE120D3-20241210DL		SDG No.:	P5261
Lab Sample ID:	P5261-19DL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	980	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035860.D	5	12/12/24 10:30	12/26/24 19:05	PB165596

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	11.4	D	0.35	1.00	1.00	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.45		30 - 150		113%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.50		30 - 150		124%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.57	*	55 - 111		142%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.45	*	53 - 106		113%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.70	*	58 - 132		175%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2550		7.257			
1146-65-2	Naphthalene-d8	5620		10.009			
15067-26-2	Acenaphthene-d10	3830		13.925			
1517-22-2	Phenanthrene-d10	8370		16.698			
1719-03-5	Chrysene-d12	5880		20.947			
1520-96-3	Perylene-d12	5420		23.029			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products